

## Calibration Certificate

Following voltages have been checked & adjusted on Blood Cell Counter Swelab Alfa Sr.No. – 15610 at Sachdeva Diagnostics, E-991, Block E, Saraswati Vihar, Pitam Pura, Delhi – 110034.

- System Voltages are in the specified range –Checked by Digital Multimeter are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Main Power Supply	219 VAC	220 +/- 5.0 VAC
2	24 Voltage Supply	24.1 VDC	24.0 +/- 1.0 VDC
3	12 Voltage Supply	12.1 VDC	12.0 +/- 0.5 VDC
4	HGB LED Voltage	4.52 VDC	4.50 +/- 0.3 VDC

- Checked the Gain and Offset of Analog Board and found OK.  
➤ All instrument thru' the Service Program and found OK.  
➤ Checked the Pumps & Valves and found OK.  
➤ Calibrate the instrument and Run control. Control results were OK  
➤ Next Calibration will be performed on 01/04/2023.

For AAMX Biomedicals,



Nitin Choudhary

Zonal Manager

Date: 02/04/2022