


Quality Control (LJ) Chart

Test	HB	Machine	SEWLAB
Control	HIGH		
Lot	B0222H		

Month	MARCH	Year	2022
-------	-------	------	------

Sl. No	Result
1	16.6
2	16.3
3	17.3
4	17.8
5	17.2
6	17.7
7	17.6
8	17.5
9	17.8
10	17.7
11	17.7
12	17.8
13	17.9
14	18.1

SD	0.50
Mean	17.50
(+) 3SD	19.01
(+) 2SD	18.51
(+) 1SD	18.00
(-) 1SD	17.00
(-) 2SD	16.49
(-) 3SD	15.99



1.25
[Signature]

Quality Control (LJ) Chart


Test	TLC	Machine	SEWLAB
Control	HIGH		
Lot	B0222H		

Month MARCH

Year 2022

SL No	Result
1	18.8
2	18
3	19.9
4	20.6
5	20.0
6	20.4
7	20.4
8	20.3
9	20.7
10	21.2
11	20.7
12	21.1
13	21.9
14	21.9

SD	1.06
Mean	20.42
(+) 3SD	23.60
(+) 2SD	22.54
(+) 1SD	21.48
(-) 1SD	19.36
(-) 2SD	18.30
(-) 3SD	17.24




17.24

Quality Control (I.J) Chart

Test: PBR Machine: SP-1000
 Control: 1000
 Lab: BR/2022

Month: MARCH Year: 2022

S/L No	Result	SD
1	5.81	1.24
2	5.72	Mean
3	5.84	5.74
4	5.1	(-1) 2SD
5	5.5	5.5
6	5.9	5.5
7	5.8	(-1) 2SD
8	5.8	5.74
9	5.5	(-1) 2SD
10	5.5	5.74
11	5.6	(-1) 2SD
12	5.9	5.98
13	6	(-1) 2SD
14	5.96	5.98



Month MARCH


Year 2022

Quality Control (LJ) Chart

Test	PCV	Machine	SEWLAB
Control	HIGH		
Lot	B0222H		

L No	Result
1	48.5
2	46.7
3	52.8
4	54.7
5	49.6
6	53.4
7	52.6
8	53.1
9	52.8
10	52.7
11	52.6
12	54.4
13	56.1
14	56.3

SD	2.71
Mean	52.59
(+) 3SD	60.71
(+) 2SD	58.00
(-) 1SD	49.89
(-) 2SD	47.18
(-) 3SD	44.47




Handwritten signature and date: 25/3

Month MARCH Year 2022

Quality Control (LJ) Chart
 Test MCV Machine SEWLAB
 Control HIGH
 Lot B0222H

SL No	Result	SD
1	89.6	1.57
2	89.4	Mean
3	90.4	91.49
4	89.4	(+) 3SD
5	90.6	96.20
6	90.7	(+) 2SD
7	91.1	94.63
8	91.9	(+) 1SD
9	92.1	93.06
10	92.1	(-) 1SD
11	93.3	89.92
12	92.6	(-) 2SD
13	93.4	88.35
14	94.3	(-) 3SD
		86.78



Handwritten signature

Month MARCH


Year 2022

Test MCH Machine SEWI,AB
 Control HICH
 Lot B022211

Quality Control (LJ) Chart

Sl. No	Result
1	30.7
2	31.2
3	29.7
4	29.2
5	31.3
6	30.1
7	30.5
8	30.4
9	31.2
10	31.0
11	31.4
12	30.3
13	29.9
14	30.4

SD	0.66
Mean	30.52
(+) 3SD	32.49
(+) 2SD	31.84
(-) 1SD	29.86
(-) 2SD	29.21
(-) 3SD	28.55



Month MARCH


Year 2022

Sl. No	Result
1	34.3
2	34.8
3	32.8
4	32.6
5	34.6
6	33.2
7	33.4
8	33.1
9	33.8
10	33.6
11	33.6
12	32.7
13	32.0
14	32.2

SD	0.85
Mean	33.34
(+) 3SD	35.89
(+) 2SD	35.04
(+) 1SD	34.19
(-) 1SD	32.48
(-) 2SD	31.63
(-) 3SD	30.78

Quality Control (LJ) Chart

Test	MIC HC	Machine	SEWLAB
Control	HIGH		
Lot	B0222H		



Handwritten signature

Quality Control (LJ) Chart

Month MARCH Year 2022

Test	PLATLET	Machine	SEWLAB
Control	HIGH		
Lot	B0222H		

Sl. No	Result
1	5.38
2	5.32
3	5.58
4	6.11
5	5.5
6	5.8
7	5.8
8	5.7
9	5.7
10	5.7
11	5.6
12	5.8
13	5.8
14	5.89

SD	0.20
Mean	5.70
(+) 3SD	6.31
(+) 2SD	6.11
(+) 1SD	5.90
(-) 1SD	5.49
(-) 2SD	5.29
(-) 3SD	5.08

