


CERTIFICATE OF VALIDATION

This is to certify that the Hematology Analyzer **Celltac ALFA, Model: MEK 7300K SERIAL No.: 994**, have been calibrated at **SDC Diagnostics Center LLP** The cell counter is successfully calibrated on 01/1/2022 with MEK-5D_PLUS218 baring expiry 30days. Control was run after calibration with MEK5D(Low,Normal,High) Lot No-B219, results found satisfactory & within specified range. The equipment is found to be working satisfactory after the calibration. The next due date of calibration is 31/12/2023, Validity +/- 30 days.

For, 
Nileshtelang

Sr.Engg-Customer Support

Nihon Kohden India Pvt. Ltd