



Lab Code No. 1097

BASIC SEROLOGY
FINAL RESULT ASSESSMENT

CYCLE NO. : 11

ROUND: 3

TOTAL PARTICIPANT: 358

DATE: 25/05/2022

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
S1- Typhoid	303	Non-Reactive	Reactive : 07 Non-Reactive : 296	98%	Within Consensus
S2- RF	263	Non-Reactive	Reactive : 09 Non-Reactive : 254	96%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

End of Report

Prepared by: ss

Page 1 of 1

Doc. No.: RML/FR/05/R-01/DL: 05.01.2022

Address: B-171, Noida Nagar, Lucknow - 226 020. Ph: 4034100-130 (30 Lines), 4077180, 2788444 Fax: 052232788555

Email: mlresearch@rml.com Website: www.rmlpathology.com





RML – Quality Assurance Program (RML – QAP)



Lab Code No.

1097

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 11

ROUND: 1

TOTAL PARTICIPANT: 294

DATE: 21/02/2022

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
S1- HIV	256	Non-Reactive	Reactive : 0 Non-Reactive : 256	100%	Within Consensus
S2- HBsAg	258	Reactive	Reactive : 248 Non-Reactive : 10	96%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

****End of Report****

Prepared by: NS

Page 1 of 1



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence



Lab Code No. 1097

BASIC SEROLOGY
FINAL RESULT ASSESSMENT

CYCLE NO. : 11

ROUND: 2

TOTAL PARTICIPANT: 328

DATE: 5/04/2022

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
S1- CRP	260	Non-Reactive	Reactive : 25 Non-Reactive : 235	90%	Within Consensus
S2- RPR	269	Reactive	Reactive : 43 Non-Reactive : 226	84%	Outside Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

End of Report

Prepared by: JP





Lab Code No. 1097

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 10

ROUND: 5

TOTAL PARTICIPANT: 288

DATE: 02/10/2021

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
CRP	236	Negative	Positive : 42 Negative : 194	82%	Within Consensus
ASO	181	Negative	Positive : 03 Negative : 178	98%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra


Programme Director

Dr. Bandana Mehrotra

Checked By:

End of Report





RML – Quality Assurance Program (RML – QAP)



Lab Code No. 1097

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 10

ROUND: 4

TOTAL PARTICIPANT: 279

DATE: 31/07/2021

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
RPR	240	Negative	Positive : 21 Negative : 219	91%	Within Consensus
Anti HCV	250	Non-Reactive	Reactive : 1 Non-Reactive : 249	99%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

End of Report


Page 1 of 1



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence



Lab Code No. 1097

BASIC SEROLOGY
FINAL RESULT ASSESSMENT

CYCLE NO. : 10

ROUND: 3

TOTAL PARTICIPANT: 279

DATE: 31/07/2021

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
Dengue IgM	245	Non-Reactive	Reactive : 4 Non-Reactive : 241	98%	Within Consensus
HBsAg	262	Non-Reactive	Reactive : 1 Non-Reactive : 261	99%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

End of Report

