

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Creatinine	Erba Chem 7	1.24	mg/dl	2.76
* Bilirubin Direct	Erba Chem 7	0.51	mg/dl	*
! Cholesterol	Erba Chem 7	117.20	mg/dl	-2.69
! HDL	Erba Chem 7	24.46	mg/dl	-2.03

Legend @ : z score ≤ 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥ 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Erba Chem 7

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	61.54	62.88	-1.33	-2.10
@ Urea	mg/dl	22.69	22.85	0.54	-0.18
! Creatinine	mg/dl	1.24	1.01	2.76	2.81
* Bilirubin Direct	mg/dl	0.51		*	5.82
@ Total Protein	g/dL	6.46	7.09	-0.56	0.09
@ Albumin	g/dL	3.42	3.74	-0.87	-0.68
@ ALP	U/L	59.43	47.98	0.84	1.31
@ GGT	U/L	23.45	23.02	-1.15	-0.34
@ Uric Acid	mg/dl	4.96	5.12	0.60	0.80
! Cholesterol	mg/dl	117.20	145.18	-2.69	-2.15
@ Triglycerides	mg/dl	124.25	109.63	1.89	1.52
! HDL	mg/dl	24.46	33.25	-2.03	-2.12
@ ALT	U/L	5.26	7.03	-1.16	-1.01
@ AST	U/L	17.21	20.20	0.25	-0.07

Legend @ : z score ≤ 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥ 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	14
Not Evaluated Parameters	0
Evaluated Parameters	14
Outlier Parameters (X)	1
EQAS Score Biochemistry	92.86 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Glucose	Erba Chem 7	219.65	mg/dl	-2.86
! Creatinine	Erba Chem 7	3.91	mg/dl	2.85
! Bilirubin Total	Erba Chem 7	0.78	mg/dl	2.46
X Bilirubin Direct	Erba Chem 7	0.88	mg/dl	5.82
! HDL	Erba Chem 7	22.03	mg/dl	-2.20
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Erba Chem 7

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! Glucose	mg/dl	219.65	248.25	-2.86	-2.10
@ Urea	mg/dl	17.52	17.66	-0.90	-0.18
! Creatinine	mg/dl	3.91	3.20	2.85	2.81
! Bilirubin Total	mg/dl	0.78	0.74	2.46	2.46
X Bilirubin Direct	mg/dl	0.88	0.40	5.82	5.82
@ Total Protein	g/dL	8.96	8.42	0.73	0.09
@ AST	U/L	14.55	20.58	-0.38	-0.07
@ ALP	U/L	55.48	42.23	1.78	1.31
@ GGT	U/L	21.15	17.78	0.47	-0.34
@ Albumin	g/dL	4.64	5.21	-0.48	-0.68
@ Uric Acid	mg/dl	4.39	4.33	0.99	0.80
@ Cholesterol	mg/dl	112.15	127.11	-1.60	-2.15
@ Triglycerides	mg/dl	113.70	100.85	1.14	1.52
! HDL	mg/dl	22.03	33.77	-2.20	-2.12
@ ALT	U/L	6.98	7.89	-0.85	-1.01

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	1
EQAS Score Biochemistry	93.33 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070