

Date: 25.02.2022

### Certificate of Calibration

**Customer Name:** Dr. MONIKA GOEL C/o.CARE HOSPITAL

**Model:** Semi-Automated Bio Chemistry Analyzer Erba Chem5 V2

**Serial No.** 90140

**Calibration done date:** 25.02.2022

**Next Calibration due Date:** 24.02.2023

**Lab In charge:** Dr. MONIKA GOEL

*This is to certify that above mentioned product has been verified of calibration  
By*

*Checking Lamp Voltage*


*Checking gain of Photo meter for all Filters*

*Checking sipping volume of peristaltic pump*

Filter	Offset	Gain
340	-3528	121884
405	-337	131262
450	-3395	144185
505	-3326	136761
546	-3237	121074
578	-3182	118150
600	-3368	119188
670	-3221	121853

*Followed by calibration of all the three types of chemistries (picked up  
CHOL, CREAT and SGPT) and QC run with satisfactory values of QC*

Calibration at site performed by Mr.DINESH SINGH

  
Authorized Signatory  
(Mr. DINESH SINGH)  
Service Engineer  
For Transasia Bio-Medicals Ltd.

**TRANSASIA BIO-MEDICALS LTD.**

Transasia House, 8 Chandivall Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax : (022) 2857 3030  
Email : transasia@transasia.co.in Website : www.transasia.co.in