

# TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C  
A NABL Accredited Laboratory

Tel. : 022-28927149  
Email : info@technical-resources.in  
technicalresources@gmail.com  
Handset : 9819103181  
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

## Calibration Certificate

<b>Calibration field:</b> Mechanical-RPM <b>Accreditation valid upto:</b> 11.04.2024 <b>Requisition no:</b> June2022-85 <b>Calibration done at</b> At Site	<b>NABL Accreditation no:</b> CC-2112 <b>Certificate no:</b> TR/2022/RPM/135 <b>ULR no:</b> CC211222000001642 <b>Date of calibration:</b> 03.06.2022 <b>Agreed Next Calib On:</b> 02.06.2023 <b>Page no:</b> 1 of 2
<b>Customer details:</b> <b>Name:</b> Shreeji Pathological Laboratory <b>Address</b> 1st Floor, Radha Vishweshwar Apartment, Narsing Lane, Opp. N L High School, Malad West, Mumbai - 400064	<b>Reference used:</b> Tachometer DT-2001B <b>Master:</b> O33442 ID No. RMOO2 <b>Master Ser. No.:</b> M-210720-7-6 Godrej CC-2634 <b>Traceable Certificate:</b> <b>Valid upto:</b> 24.07.2022
<b>Details of DUC:</b> <b>Equipment:</b> Centrifuge <b>Make:</b> Remi <b>Model:</b> R8C DX <b>Serial No.</b> FBLC-5760 <b>ID No.</b> Not Specified <b>Location</b> Pathology Lab	<b>Condition of item:</b> Good <b>Range RPM:</b> Not Specified <b>Accuracy:</b> Not Specified <b>Resolution (RPM):</b> 1
<b>Ambient Temp(°C):</b> 25+/- 2 <b>Humidity Rh(%):</b> 50 +/-20	

**Remark:**

1. Calibration procedure used: LAB-SOP/08
2. The Reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approx. 95%.
3. The observations reported represent values at the time of measurement. They do not convey any long term stability information.
4. The standards maintained are traceable to National / International Standards through accredited laboratories.
5. Sticker affixed on the unit

**Note:**

1. This certificate refer only to particular instrument submitted for calibration.
2. This certificate shall not be reproduced.
3. Technical Resources shall not be responsible for any change in calibration date and performance specification on account of malfunctioning of the instrument due to damage caused to it after handing over it to the customer.
4. Any queries for the calibration certificate will be clarified only if raised with in 30 days from date of issue. The reissue of certificate after 30 days will be chargeable.

Authorised signature  
Sr Technical Manager  
N Srinivas Rao



# TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.  
A MSME Registered Unit, NSIC-CRISIL Rated SE 3C  
A NABL Accredited Laboratory

Tel. : 022-28927149  
Email : info@technical-resources.in  
technicalresources@gmail.com  
Handset : 9819103181  
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

## Calibration data:

Equipment:	Centrifuge	Certificate no:	TR/2022/RPM/135
Serial no:	FBLC-5760	ULR no:	CC211222000001642
ID no.	Not Specified	Date of calibration:	03.06.2022
		Page No:	2 of 2

## RUN DETAILS

Set RPM Display	Avg. Reading on UUC Display RPM	Avg. Reading On Ref. RPM	Deviation in RPM (+/-)	Expanded Uncertainty in RPM (+/-)
1000	1000.0	1005.6	-5.6	25.0
2500	2500.0	2504.3	-4.3	25.0
4000	4000.0	4006.3	-6.3	25.0

END OF RESULT

04.06.2022

Certificate Issue Date:

Checked By

*Dilip*

Authorised signatory  
Sr Technical Manager  
N Srinivas Rao



F7.8.4

**CAL RESOURCES**

ANALYTICAL SERVICE FRANCHISE PROGRAM  
ANALYTICAL & HISTORICAL LABORATORY COMPANY  
A Registered Unit, NCEM, RUC 0005 10

Temp 20-29.6  
On 03.06.21  
01.06.20  
info@calresources.com