



CERTIFICATE OF CALIBRATION

Instrument: Mispa CXL PRO Fully Automated Biochemistry Analyser

Customer Name: Sevana Hi-tech Lab, Parappur, Thrissur

Date of Installation : 09/11/2021

S/N: 210T180CS0345K

Alignment Check

Motion Check	Checked	Remark
Sample Probe	<input checked="" type="checkbox"/>	OK
Reagent Probe	<input checked="" type="checkbox"/>	OK
Mixing Bar	<input checked="" type="checkbox"/>	OK
Reaction Disk	<input checked="" type="checkbox"/>	OK
Reagent Disk	<input checked="" type="checkbox"/>	OK
Sample Disk	<input checked="" type="checkbox"/>	OK
Cuvette Check	<input checked="" type="checkbox"/>	OK
Lamp	<input checked="" type="checkbox"/>	OK

System	Checked	Remark
Main unit	<input checked="" type="checkbox"/>	OK
Temp. unit	<input checked="" type="checkbox"/>	OK
Mixing unit	<input checked="" type="checkbox"/>	OK
Reaction unit	<input checked="" type="checkbox"/>	OK
Reagent unit	<input checked="" type="checkbox"/>	OK
Sample unit	<input checked="" type="checkbox"/>	OK

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,
 Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4 300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in

OPTICAL SECTION

Filter	Light Intensity	
340	OK	
380	OK	
405	OK	
450	OK	
480	OK	
505	OK	
546	OK	
570	OK	
600	OK	
660	OK	
700	OK	
800	OK	

Power Supply : 12V 24V

The equipment is calibrated for the different parameters by using Multi Calibrator

Lot.No: 31110054 Exp: 2023-10

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the Agappe Diagnostics.

Date of Calibration : **09/11/2021**

Next Due : **08/05/2022**

**For, Agappe Diagnostics Ltd,
Authorised signatory**


Name: Nimmy P R

Designation: Product Specialist

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PL115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/SBS-CB Block No. 32, KINFRA Small Industrial Park,
Nellai, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4 300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shrawji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in