

Result Reprint

Report Type

: Controls

09-Jun-2022

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD
1	26491	BIROAD L1	GPT1	27.1 U/L		155591	09-Jun-2022 12:48:11	27.400	3.600
2	26491	BIROAD L1	BIT1	1.03 mg/dl		155597	09-Jun-2022 12:49:59	0.990	0.160
3	26491	BIROAD L1	GLU1	80.6 mg/dl	F,-1SD	155599	09-Jun-2022 12:50:35	91.700	9.300
4	26491	BIROAD L1	UA1	5.4 mg/dl		155600	09-Jun-2022 12:50:53	5.380	0.540
5	26491	BIROAD L1	CHOL1	251 mg/dl		155601	09-Jun-2022 12:51:11	251.000	26.000
6	26491	BIROAD L1	HDLC1	60.0 mg/dl		155602	09-Jun-2022 12:51:29	67.200	9.000
7	26491	BIROAD L1	GOT2	38.2 U/L		155608	09-Jun-2022 12:53:17	37.100	5.000
8	26491	BIROAD L1	UREA1	33.6 mg/dl		155639	09-Jun-2022 14:19:11	33.800	3.400
9	26491	BIROAD L1	TRIG1	181.8 mg/dl		155656	09-Jun-2022 14:24:18	182.000	18.000
10	26491	BIROAD L1	PRO1	6.83 g/dl		155661	09-Jun-2022 14:25:48	6.510	0.660
11	26491	BIROAD L1	ALB1	4.08 g/dl		155666	09-Jun-2022 14:27:19	3.960	0.300
12	26491	BIROAD L1	ALP2	120 U/L		155686	09-Jun-2022 15:28:43	115.000	16.000
13	26491	BIROAD L1	CRE1	1.65 mg/dl		155688	09-Jun-2022 16:06:54	1.660	0.220
14	26491	BIROAD L1	CA1	9.6 mg/dl		155691	09-Jun-2022 16:07:48	9.210	0.930