

## VIEW LAB MONTHLY SUMMARY

Lab Name	CARE HOSPITAL	Details About Robust Analysis
Lab No	9041	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2022	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 05/04/2022

Date of Report Published : 05/05/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	799	223.18	10.20	22.77	210 mg/dL	-0.58	1.61
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	784	38.76	12.99	5.03	42.62 mg/dL	0.77	0.36
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	518	4.68	10.61	0.50	4.28 mg/dL	-0.80	0.04
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	703	2.90	11.57	0.34	1.87 mg/dL	-3.03	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	672	5.80	8.17	0.47	5.84 g/dL	0.08	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	649	3.38	8.14	0.28	2.97 g/dL	-1.49	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	730	7.12	12.13	0.86	7.06 mg/dL	-0.07	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	754	111.29	11.48	12.78	97 mg/dL	-1.12	0.93
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	726	114.85	16.13	18.52	119.7 mg/dL	0.26	1.37

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

VIEW LAB MONTHLY SUMMARY

Lab Name	CARE HOSPITAL	Details About Robust Analysis
Lab No	9041	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2022	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 04/03/2022

Date of Report Published : 06/04/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	109.92	9.69	10.65	104 mg/dL	-0.56	0.76
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	741	78.61	11.58	9.11	79.42 mg/dL	0.09	0.67
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	459	3.84	11.18	0.43	3.59 mg/dL	-0.58	0.04
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	668	1.57	19.28	0.30	1.46 mg/dL	-0.36	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	593	5.34	8.62	0.46	8.42 g/dL	6.7	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	633	3.15	9.30	0.29	3.62 g/dL	1.60	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	637	8.01	11.65	0.93	7.51 mg/dL	-0.54	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	700	105.41	11.28	11.89	101 mg/dL	-0.37	0.90
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	679	188.91	11.47	21.66	186 mg/dL	-0.13	1.66

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

VIEW LAB MONTHLY SUMMARY

Lab Name	CARE HOSPITAL	Details About Robust Analysis
Lab No	9041	Detail About Monthly Summary
Month	February	Detail about SDI
Year	2022	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 08/02/2022

Date of Report Published : 05/03/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	728	278.22	10.88	30.27	253.8 mg/dL	-0.81	2.24
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	774	31.02	13.91	4.32	36.2 mg/dL	1.20	0.31
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	480	1.14	12.72	0.14	1.38 mg/dL	1.66	0.01
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	614	5.06	11.12	0.56	3.6 mg/dL	-2.59	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	604	5.81	9.02	0.52	6.16 g/dL	0.67	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	597	3.40	8.90	0.30	3.5 g/dL	0.33	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	524	4.45	12.44	0.55	5.01 mg/dL	1.01	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	678	112.37	11.73	13.18	108 mg/dL	-0.33	1.01
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	661	88.35	21.56	19.05	87.6 mg/dL	-0.04	1.48

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## VIEW LAB MONTHLY SUMMARY

Lab Name	CARE HOSPITAL	Details About Robust Analysis
Lab No	9041	Detail About Monthly Summary
Month	January	Detail about SDI
Year	2022	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/01/2022

Date of Report Published : 07/02/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	687	130.44	8.67	11.30	133.1 mg/dl	0.24	0.86
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	740	47.34	12.33	5.84	54.2 mg/dl	1.18	0.43
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	434	1.70	10.33	0.18	1.72 mg/dl	0.11	0.02
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	615	2.30	15.28	0.35	2.01 mg/dl	-0.83	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	556	5.38	9.07	0.49	5.82 g/dl	0.90	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	577	3.15	8.76	0.28	2.87 g/dl	-1.01	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	592	7.17	12.44	0.89	7.1 mg/dl	-0.08	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	636	101.83	11.62	11.83	100.3 mg/dl	-0.13	0.94
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	577	115.89	15.40	17.85	103.2 mg/dl	-0.71	1.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

VIEW LAB MONTHLY SUMMARY

Lab Name	CARE HOSPITAL	Details About Robust Analysis
Lab No	9041	Detail About Monthly Summary
Month	December	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

All Analyser Result
Print

Date of Result Entered : 11/12/2021

Date of Report Published : 03/01/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	735	220.59	10.06	22.19	175.8 mg/dl	-2.02	1.64
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	762	105.93	13.00	13.77	109.1 mg/dl	0.23	1.00
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	488	2.27	9.44	0.21	2.22 mg/dl	-0.23	0.02
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	649	2.93	11.30	0.33	3.34 mg/dl	1.24	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	665	5.34	9.22	0.49	5.7 g/dl	0.73	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	650	3.21	9.20	0.30	3.13 g/dl	-0.27	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	637	4.08	14.87	0.61	4.65 mg/dl	0.94	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	716	104.99	12.01	12.61	99.78 mg/dl	-0.41	0.94
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	605	85.49	17.53	14.98	180 mg/dl	6.31	1.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## LAB MONTHLY SUMMARY



Lab Name CARE HOSPITAL  
 Month November  
 Constituent Group Chemistry II

Lab No 9041  
 Year 2021



Date of Result Entered : 17/11/2021

Date of Report Published : 03/12/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	775	144.88	10.50	15.21	141 mg/dl	-0.26	1.09
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	844	43.73	12.85	5.62	44 mg/dl	0.05	-0.39
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	484	9.03	13.47	1.22	8.56 mg/dl	-0.39	0.11
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	699	2.27	16.68	0.38	2.1 mg/dl	-0.45	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	624	5.52	8.94	0.49	5.94 g/dl	0.85	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	634	3.25	9.15	0.30	3.22 g/dl	-0.10	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	665	8.17	14.31	1.17	8.28 mg/dl	0.09	0.09
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	734	108.76	10.65	11.58	135.1 mg/dl	2.27	0.85
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	675	145.97	13.24	19.32	183 mg/dl	1.92	1.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

## LAB MONTHLY SUMMARY



Lab Name CARE HOSPITAL

Lab No 9041

Month October

Year 2021

Constituent Group Chemistry II



Date of Result Entered : 12/10/2021

Date of Report Published : 01/11/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	803	105.70	10.78	11.39	90.1 mg/dl	-1.37	0.80
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	887	29.20	15.70	4.58	30.34 mg/dl	0.25	0.31
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	488	1.31	11.94	0.16	1.46 mg/dl	0.96	0.01
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	692	6.97	13.68	0.95	5.9 mg/dl	-1.12	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	637	5.42	10.60	0.57	5.51 g/dl	0.16	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	636	3.21	9.77	0.31	3.17 g/dl	-0.13	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	689	5.97	13.21	0.79	6.23 mg/dl	0.33	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	651	101.83	11.23	11.43	157.1 mg/dl	4.84	0.90
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	570	92.74	16.83	15.61	114.4 mg/dl	1.39	1.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
CMC EQAS Co-Ordinator  
Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

LAB MONTHLY SUMMARY



Lab Name CARE HOSPITAL  
 Month September  
 Constituent Group Chemistry II

Lab No 9041  
 Year 2021



Date of Result Entered : 07/09/2021

Date of Report Published : 05/10/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	770	399.21	11.44	45.67	423 mg/dl	0.52	-3.29
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	820	86.31	15.24	13.16	85.01 mg/dl	-0.10	0.92
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	483	3.37	11.05	0.37	3.57 mg/dl	0.54	0.03
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	670	2.10	15.07	0.32	2.87 mg/dl	2.43	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	597	5.39	8.88	0.48	5.53 g/dl	0.29	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	616	3.22	9.51	0.31	3.17 g/dl	-0.16	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	675	8.55	14.14	1.21	9.35 mg/dl	0.66	0.09
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	713	112.13	10.78	12.09	100.5 mg/dl	-0.96	0.91
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	658	217.36	11.08	24.09	293.7 mg/dl	3.17	1.88

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*



## LAB MONTHLY SUMMARY



Lab Name CARE HOSPITAL  
 Month August  
 Constituent Group Chemistry II

Lab No 9041  
 Year 2021



Date of Result Entered : 17/08/2021

Date of Report Published : 01/09/2021

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	691	253.55	11.35	28.78	272 mg/dl	0.64	2.19
2	UREA II	Others ( Urease Berthelot / Nod Dye ) II	Any Analyser (Automation / Semi Automation)	873	44.63	13.73	6.13	41.4 mg/dl	-0.53	0.41
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	472	1.96	12.60	0.25	1.97 mg/dl	0.04	0.02
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	618	3.97	11.54	0.46	2.89 mg/dl	-2.36	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	619	5.52	9.17	0.51	6.04 g/dl	1.03	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	594	3.29	8.83	0.29	3.39 g/dl	0.34	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	643	3.54	22.19	0.78	5 mg/dl	1.86	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	674	107.45	11.66	12.52	90.4 mg/dl	-1.36	0.96
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	629	120.29	15.63	18.81	136.2 mg/dl	0.85	1.50

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

LAB MONTHLY SUMMARY



Lab Name CARE HOSPITAL  
 Month July  
 Constituent Group Chemistry II

Lab No 9041  
 Year 2021



Date of Result Entered : 19/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SD	SI	II
						CV	SD				
1	GLUCOSE II	GOD-POD II	Transasia / Erba	751	148.33	10.85	16.09	134.9 mg/dl	-0.83	1.17	
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	853	34.55	15.04	5.20	33.58 mg/dl	-0.19	0.36	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	479	2.77	11.56	0.32	2.28 mg/dl	-1.53	0.03	
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	655	2.84	12.80	0.36	2.27 mg/dl	-1.57	0.03	
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	609	5.55	8.57	0.48	5.66 g/dl	0.23	0.04	
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	607	3.32	8.78	0.29	2.84 g/dl	-1.65	0.02	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	686	4.95	17.29	0.86	6.04 mg/dl	1.27	0.07	
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	783	112.10	13.65	15.31	128 mg/dl	1.04	1.09	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	666	167.18	13.26	22.17	166.16 mg/dl	-0.05	1.72	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*