

Parashar Micro Measurement Pvt. Ltd.

Unit-II : J-22, Site-C, Surajpur Industrial Area, UPSIDC, Greater Noida (U.P.)

PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com

Website : www.pmmpl.in, www.parashar.info



The way to Technology
(17 Years of Excellence)


Calibration Certificate

						Page 1
Certificate Number : PMM/BM/731437						
Calibrated For		M/s. Som Datt Medical Centre S - 524, School Block , Shakarpur Delhi - 110092		Calibrated on		31.01.2022
				Valid Upto		30.01.2023
				Cert. Issue Date		01.02.2022
Description of Equipments						
Nomenclature		Bio Chemistry Analyzer		Accuracy/Class		As per specification
Test Type		Electrical Safety Test		Condition of UUC		Physically Ok
Make		Mispa ACE		Party ID Mark No.		9
Model No.		AGAPPE		Location		Path Lab
Mfr. S.No.		WK-671048277				
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Electrical Safety Analyzer	Fluke/ESA612	PMM/BM/ESA-01	FLUKE CORPORATION	PC-EVT-0D3B	31.03.2022
Standard used for calibration are traceable to National/International Standards through unbroken chain of calibration.						
Calibration Performed at		AT SITE		Environmental		Temperature = 25 ± 5° C
Calibration Procedure No.		PMM/II/BMCP/01		Condition(s)		Humidity = 50 ± 10 % RH
Reference Standard / Guideline		IEC-60601-1				
CALIBRATION RESULTS						
1.0 Main Line Test						
a. Main Line Voltage supply working properly						Yes
b. Main Line ground connected properly						Yes
2.0 Earth Resistance Test						
Sr.No.	Max Limit (Ω)	Actual Observed Value (Ω)		Remark		
1	2.0	0.229		Pass		
3.0 Polarity Test Pass						Yes
4.0 Safety test						
Sr.No.	Parameter	Permitted Values	Measured Values	Deviation%	Remarks (Pass/Fail)	
1	Ground Wire Resistance	<0.3Ω	0.251	-0.00049	Pass	
2	Chassis Leakage	<100 μANC	96μA	-0.04	Pass	
		<500μASFC	N/A	N/A	N/A	
3	Pateint Leakage Current	<100μABandB F	N/A	N/A	N/A	
		<10μACF	1.3μA	-0.087	Pass	
4	Patient Lead Leakage Current, Isolation Test	<100μABF	N/A	N/A	N/A	
		<10μACF	1.2μA	-0.088	Pass	
5	Insulation (Optional) 500V	<2MΩ	OR MΩ	N/A	N/A	

Uncertainty of measurement (at approx. 95% confidence level with coverage factor k=2)

1.23%

Calibrated By : Gordhan Baurwa
Designation : Calibration Engineer



Approved By : Vikram Parashar
Designation : Technical Manager

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.



INSTALLATION QUALIFICATION

For

FULLY AUTOMATED CLINICAL CHEMISTRY ANALYZER

MODEL : MISPA ACE

AGAPPE DIAGNOSTICS LTD.

Stamp
28/5/22

ADL/DO/SER/IOP/0011/22-23

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091. -
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**TABLE OF CONTENTS**

S No.	Contents	Page No.
I	Approval of the IQ procedure	3
II	Instructions	4
III	Scope	5
V	Installation Qualification	6
V	Comments	11
VI	System Certification	12

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413
CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)

401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

S-6, IInd Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



AGAPPE

I. APPROVAL OF THE INSTALLATION QUALIFICATION PROCEDURE

Both **SOM DATT MEDICAL CENTRE, NEW DELHI** and **AGAPPE DIAGNOSTICS LTD.** are jointly responsible for the installation of **MISPA ACE** with serial number **WK-67104827T** at **"Som Datt Medical Centre, Poly Clinic Cum Diagnostic Centre, S-524, School Block, New Delhi"** as per the attached Installation Qualification protocol.

Protocol Performed By : Representative of AGAPPE DIAGNOSTICS LTD.



Name	:	Vijender Pratap Singh	Signature :
Designation	:	Engineer - Customer Service	Date : 18/May/2017
Company	:	AGAPPE DIAGNOSTICS LTD.	

Validation Team from SOM DATT MEDICAL CENTRE, NEW DELHI

Jaya

Name	:	JAYAUCCAS	Signature :
Designation	:	Lab Technician	Date : 28/5/22
Department	:	Path lab	

Customer Authorizations:

Name : JAYAUCCAS
Title : INSTALLATION QUALIFICATION
Site :

Signature : *Jaya*
Date : 28/5/22

Jaya
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



II. INSTRUCTIONS

1. An authorized **AGAPPE DIAGNOSTICS LTD.** representative will check and enter the specific data as outlined in the Installation Qualification. Each result will be noted and dated.
2. The concerned employees of **SOM DATT MEDICAL CENTRE, NEW DELHI** will verify each result and sign in each page. The member of the validation team will carry this out.
3. ALL deviations from the acceptance criteria detailed in this document will be noted in the **COMMENTS** section at the end of the IQ protocol. All resolution to such problems will also be noted in the **COMMENTS** section, and must be resolved prior to issuance of a **SYSTEM CERTIFICATION**.

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091,
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]

401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE

5-6, IInd Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



III. SCOPE

This Installation Qualification protocol will be performed on the **MISPA ACE** at **SOM DATT MEDICAL CENTRE, NEW DELHI**. This Installation protocol will define the documentation that will be used to evaluate the instruments installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage.

Trained, knowledgeable personnel will perform qualification studies. Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattinattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]

401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



IV. INSTALLATION QUALIFICATION

A. Instrument Identification

Verified Date : 18/May/2017

1. Model Name : MISPA ACE
2. Serial Number : WK-67104827T

B. Following is a list of accessories /consumables verified:

S No.	Component	Present (Yes/No)	Verified By	Comments
01	Analyzer Main unit	Yes	Vijender Pratap Singh	
02	Computer	Yes	Vijender Pratap Singh	
03	Printer	Yes	Vijender Pratap Singh	
04	RS232 Connector cable	Yes	Vijender Pratap Singh	
05	Power cord	Yes	Vijender Pratap Singh	
06	Waste can	Yes	Vijender Pratap Singh	
07	Water can	Yes	Vijender Pratap Singh	
08	Water connector tube	Yes	Vijender Pratap Singh	
09	Waste connector tube	Yes	Vijender Pratap Singh	
10	Water pre filter	Yes	Vijender Pratap Singh	
11	Probe assembly	Yes	Vijender Pratap Singh	
12	Mixer assembly	Yes	Vijender Pratap Singh	
13	Reagent tray	Yes	Vijender Pratap Singh	
14	Cuvette blocks	Yes	Vijender Pratap Singh	
15	Reagent bottles	Yes	Vijender Pratap Singh	

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



C. Installation Checks

Purpose: To make sure that the instrument is received with all the major components necessary for operation

S No.	Component	(Yes/No)	Verified By	Comments
1	Major components and accessories are present	Yes	Vijender Pratap Singh	
2	There is no physical damage to the components	Yes	Vijender Pratap Singh	
3	Ambient temperature is available	Yes	Vijender Pratap Singh	
4	Instrument is installed in the table leveled properly	Yes	Vijender Pratap Singh	
5	Power supply is correct	Yes	Vijender Pratap Singh	
6	Electric connections are tight, weatherproof & earthed	Yes	Vijender Pratap Singh	
7	Analyzer unit installed as per the manufacturers recommendations	Yes	Vijender Pratap Singh	

Summary:

The information is recorded against each data and the deviations if any is justified / explained properly.

Acceptance criteria:

PARAMETER

PASS

FAIL

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/585-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**D. Instrument Installation**

Date of Installation : 18/May/2017

Installed by : Vijender Pratap Singh

S No.	Parameter	Done by	Comments
01	Installation of the main unit	Vijender Pratap Singh	
02	Installation of the computer system	Vijender Pratap Singh	
03	Installation of the printer	Vijender Pratap Singh	
04	Connecting the waste tubing	Vijender Pratap Singh	
05	Connecting the system water tubing	Vijender Pratap Singh	
06	Installing the reagent tray	Vijender Pratap Singh	
07	Installation of the probe	Vijender Pratap Singh	
08	Installation of the mixer	Vijender Pratap Singh	
09	Installing the cuvettes	Vijender Pratap Singh	

Acceptance criteria : System should be "Ready" after daily maintenance without any error

PARAMETER PASS FAIL

Parameter values for verification : System found "Ready" after daily maintenance

Top
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PD), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nelladi, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**E. Cell Blank**

Purpose: To check the cuvette blank OD at different wavelength

Procedure:

S No.	Activity	Done By	Date
01	Checking the Cell blank	Vijender Pratap Singh	18/May/2017
02	340 nm	Vijender Pratap Singh	18/May/2017
03	405 nm	Vijender Pratap Singh	18/May/2017
04	450 nm	Vijender Pratap Singh	18/May/2017
05	510 nm	Vijender Pratap Singh	18/May/2017
06	546 nm	Vijender Pratap Singh	18/May/2017
07	578 nm	Vijender Pratap Singh	18/May/2017
08	630 nm	Vijender Pratap Singh	18/May/2017
09	670 nm	Vijender Pratap Singh	18/May/2017

Acceptance criteria:

- No error Messages displayed
- All Cuvettes should show acceptance

PARAMETER **PASS** **FAIL**

Parameter values for verification: No Error Messages

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Plattinattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Wellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkattoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**F. Priming**

Purpose: To check the internal tubing's and flow of liquids

S No	Activity	Done By	Date
01	Do wash using the maintenance screen	Vijender Pratap Singh	18/May/2017

Acceptance criteria:

- No error message or air bubbles in the internal tubing's
- All the dilutor working well without air bubbles / leakage

PARAMETER PASS FAIL

Parameter values for verification: No Error Messages

Days
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413
CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PD), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park, Neilad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

BANGALORE OFFICE

5-6, 1st Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



V. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outline in this Installation Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for Operational Qualification.

Report Performed By: **AGAPPE DIAGNOSTICS LTD.** Representative

Name : Vijender Pratap Singh

Designation : Engineer - Customer Service

Company : AGAPPE DIAGNOSTICS LTD.

Signature :

Date : 18/May/2017



Customer Authorizations: **SOM DATT MEDICAL CENTRE, NEW DELHI**

Name : JAYAULLAS

Designation : Lab Technician

Title : INSTALLATION QUALIFICATION

Signature : Jays

Date : 28/5/22

Jays

Jays

AGAPPE DIAGNOSTICS LTD. | ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
K/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



OPERATIONAL QUALIFICATION

For

FULLY AUTOMATED CLINICAL CHEMISTRY ANALYZER

MODEL : MISPA ACE

AGAPPE DIAGNOSTICS LTD.

Jays
28/3/22

ADL/DO/SER/IOP/0011/22-23

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**TABLE OF CONTENTS**

S No.	Contents	Page No.
I	Approval of the OQ procedure	3
II	Instructions	4
III	Scope	5
IV	Operational Qualification	6
V	Comments	13
VI	System Certification	14

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413
CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nelload, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



I. APPROVAL OF THE OPERATIONAL QUALIFICATION PROCEDURE

Both **SOM DATT MEDICAL CENTRE, NEW DELHI** and **AGAPPE DIAGNOSTICS LTD.** are jointly responsible for the installation of **MISPA ACE** with serial number **WK-67104827T** at "**Som Datt Medical Centre, Poly Clinic Cum Diagnostic Centre, S-524, School Block, New Delhi**" as per the attached OPERATIONAL QUALIFICATION protocol.

Protocol Performed By: Representative of **AGAPPE DIAGNOSTICS LTD.**



Name	: Vijender Pratap Singh	Signature :
Designation	: Engineer - Customer Service	Date : 18/May/2017
Company	: AGAPPE DIAGNOSTICS LTD.	

Validation Team from **SOM DATT MEDICAL CENTRE, NEW DELHI**

Jaya

Name	: <i>Jayavillas</i>	Signature : <i>Jaya</i>
Designation	: <i>Lab Technician</i>	Date : <i>28/5/22</i>
Department	: <i>Path Lab</i>	

Customer Authorizations:

Name : *Jayavillas*

Title : OPERATIONAL QUALIFICATION

Signature :

Jaya

Site :

Date

28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nelladi, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



II. INSTRUCTIONS

1. An authorized **AGAPPE DIAGNOSTICS LTD** representative will check and enter the specific data as outlined in the **OPERATIONAL QUALIFICATION**. Each result will be noted and dated.
2. ALL deviations from the acceptance criteria detailed in this document will be noted in the **COMMENTS** section at the end of the OQ protocol. All resolution to such problems will also be noted in the **COMMENTS** section, and must be resolved prior to issuance of a **SYSTEM CERTIFICATION**.

Jayg
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkattaoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



III. SCOPE

This Operational Qualification protocol will be performed on the **MISPA ACE** at **SOM DATT MEDICAL CENTRE, NEW DELHI**. This Installation protocol will define the documentation that will be used to evaluate the instruments installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage. Trained, knowledgeable personnel will perform qualification studies. Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



IV. OPERATIONAL QUALIFICATION

A. Instrument Identification

Verified Date : 18/May/2017

1. Model Name : MISPA ACE
2. Serial Number : WK-67104827T

B. Following is a list of tests to be performed and verified:

Test No.	Test Name	Test purpose	Verified By and date
01	Start up	To make the equipment ready for operation	Vijender Pratap Singh 18/May/2017
02	Daily maintenance	To clean appropriate modules	Vijender Pratap Singh 18/May/2017
03	Calibration	To calibrate the system	Vijender Pratap Singh 18/May/2017
04	QC check	To confirm that the system is calibrated and working within specifications	Vijender Pratap Singh 18/May/2017
05	Reproducibility check	To check the precision [CV %] after calibration	Vijender Pratap Singh 18/May/2017
06	Sample programming and Analysis	To run the samples	Vijender Pratap Singh 18/May/2017
07	Shut down procedure	To shut down the system	Vijender Pratap Singh 18/May/2017

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/586-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaishingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



Test 01: Starting the system

Purpose: To make the instrument READY for operation

Summary:

Instrument checks functioning of different parts of the instrument automatically. If there is an error code, initialize the system and follow corrective action instructions provided for the error message.

Procedure:

- Wait for the instrument to get ready after initialization
- Check the room temperature and switch on the Air Conditioner
- Check the UPS.
- Switch on the MISPA ACE by pressing the main switch, then switch on the computer and then the system power
- Double click the analyzer icon to initialize the system software
- Input the password and do as prompted, the system menu pops up if start up is finished
- If not, initialize by again after solving the error displayed
- Follow instructions provided for the error message

Acceptance criteria: System to display READY status

PARAMETER	PASS	FAIL
-----------	------	------

Parameter values for verification: "READY" on Status Area

Jaya
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/585-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

Test 02: Daily Maintenance

Purpose: To clean appropriate modules so as per the daily maintenance protocol

Method:

Refer detailed procedure for Daily Maintenance

S No.	Activity	Done by	Date
01	Empty waste container	Vijender Pratap Singh	18/May/2017
02	Check Wash solutions	Vijender Pratap Singh	18/May/2017
03	Check system water	Vijender Pratap Singh	18/May/2017
04	Check Reagents	Vijender Pratap Singh	18/May/2017
05	Clean Reagent Block	Vijender Pratap Singh	18/May/2017
06	Clean out side area	Vijender Pratap Singh	18/May/2017
07	Load supplies and remove outdated and empty reagents	Vijender Pratap Singh	18/May/2017
08	Check main menu screen	Vijender Pratap Singh	18/May/2017
09	Perform Quality Control	Vijender Pratap Singh	18/May/2017

Acceptance criteria: System should be "Ready" after daily maintenance without any error

PARAMETER **PASS** **FAIL**

Parameter values for verification: System found "Ready" after daily maintenance

Jayg
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4380 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/585-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4358 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 03: Calibration of the system****Purpose:** To calibrate the system**Procedure:**

S No	Activity	Done By	Date
01	Preparation of the cal material	Vijender Pratap Singh	18/May/2017
02	Performing Calibration with calibration programming screen	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Calibration data shows concordance OD values in triplicate measurement

PARAMETER	PASS	FAIL
------------------	-------------	-------------

Parameter values for verification	:	No error message in the printout
--	---	----------------------------------

Test 04: QC check**Purpose:** To confirm that systems, reagents and consumables are acceptable & working within specifications for each assay used.**Procedure:**

S No.	Activity	Done By	Date
01	Preparing Biochemistry control material	Vijender Pratap Singh	18/May/2017
02	Creating QC file	Vijender Pratap Singh	18/May/2017
03	QC sample programming and analysis	Vijender Pratap Singh	18/May/2017

Acceptance criteria: QC results within specified limits mentioned on the control product insert

PARAMETER	PASS	FAIL
------------------	-------------	-------------

Parameter values for verification: QC values within \pm 2SD

Jayg
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 05: Reproducibility Check****Purpose:** To check the reproducibility of instrument

S No.	Activity	Done By	Date
01	Preparing control material/ sample	Vijender Pratap Singh	18/May/2017
02	Running triplicates	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Results CV within specified limits

PARAMETER	PASS	FAIL
-----------	------	------

Parameter values for verification: CV% within the limits**Test 06: Sample programming and Analysis****Purpose:** To run the samples**Procedure:**

S No.	Activity	Done By	Date
01	Preparing and Processing of samples	Vijender Pratap Singh	18/May/2017
02	Programming samples	Vijender Pratap Singh	18/May/2017
03	Aspirating the samples	Vijender Pratap Singh	18/May/2017
04	Viewing samples in process	Vijender Pratap Singh	18/May/2017
05	Review results: Monitoring results	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Samples Analysis without any error

PARAMETER	PASS	FAIL
-----------	------	------

Parameter values for verification: Sample analysis without any error

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 07: Shut down****Purpose:** To shut down the system**Procedure:**

S No.	Activity	Done By	Date
01	Preparing the system for shut down.	Vijender Pratap Singh	18/May/2017
02	Following the on screen instructions	Vijender Pratap Singh	18/May/2017
03	Switch off the instrument when prompted	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Shut down without any error

PARAMETER	PASS	FAIL
-----------	------	------

Parameter values for verification: Shut down without any error

V. OPERATIONAL PROCEDURE**Certificate of Training****1. Technician Training**

This certifies that the technicians have received basic user training in the following categories for the system described in this Operational Qualification.

S No.	Training program	Initials	Date
1	Instrument Setup	Vijender Pratap Singh	18/May/2017
2	System Operation	Vijender Pratap Singh	18/May/2017
3	Basic trouble shooting and Maintenance	Vijender Pratap Singh	18/May/2017

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
D5M 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



2. Operator Training

The users responsible for the operation of this equipment have been trained in the proper usage of the system. Training focused on the basic operation and maintenance of the system.

S No.	Operators	Department	Initials	Date

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/585-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



VI. COMMENTS

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



AGAPPE

VII. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outline in this Operational Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for Performance Qualification.

Report Performed By: **AGAPPE DIAGNOSTICS LTD.** Representative

Name : Vijender Pratap Singh

Designation : Engineer – Customer Service

Company : AGAPPE DIAGNOSTICS LTD.

Signature :

Date :

18/May/2017



Customer Authorizations: **SOM DATT MEDICAL CENTRE, NEW DELHI**

Name : Jays

Designation : Lab Technician

Title : OPERATIONAL QUALIFICATION

Signature :

Date :

28/5/22

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. | ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



PERFORMANCE QUALIFICATION

For

FULLY AUTOMATED CLINICAL CHEMISTRY ANALYZER

MODEL : MISPA ACE

AGAPPE DIAGNOSTICS LTD.

Jays
28/5/22

ADL/DO/SER/IOP/0011/22-23

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellikud, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

S-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**TABLE OF CONTENTS**

S No.	Contents	Page No
I	Approval of the PQ procedure	3
II	Instructions	4
III	Scope	5
IV	Performance Qualification	6
V	Comments	15
VI	System Certification	16

Jays
28/3/22

AGAPPE DIAGNOSTICS LTD. | ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413
CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Emakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]

401 & 402, 4th Floor, Jaishigh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/589-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



I. APPROVAL OF THE PERFORMANCE QUALIFICATION PROCEDURE

Both **SOM DATT MEDICAL CENTRE, NEW DELHI** and **AGAPPE DIAGNOSTICS LTD.** are jointly responsible for the installation of **MISPA ACE** with serial number **WK-67104827T** at "**Som Datt Medical Centre, Poly Clinic Cum Diagnostic Centre, S-524, School Block, New Delhi**" as per the attached PERFORMANCE QUALIFICATION protocol.

Protocol Performed By: Representative of AGAPPE DIAGNOSTICS LTD.



Name	: Vijender Pratap Singh	Signature :
Designation	: Engineer - Customer Service	Date : 18/May/2017
Company	: AGAPPE DIAGNOSTICS LTD.	

Validation Team from SOM DATT MEDICAL CENTRE, NEW DELHI

Name	: JAYA ULLAS	Signature : Jaya
Designation	: Lab Technician	Date : 28/5/22
Department	: Path Lab	

Customer Authorizations:

Name : JAYA ULLAS

Title : PERFORMANCE QUALIFICATION

Site :

Signature : Jaya

Date : 28/5/22

28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



II. INSTRUCTIONS

1. An authorized **AGAPPE DIAGNOSTICS LTD.** representative will check and enter the specific data as outlined in the Performance Qualification. Each result will be noted and dated.
2. The concerned employees of **SOM DATT MEDICAL CENTRE, NEW DELHI** will verify each result and sign in each page. The member of the validation team will carry this out.
3. All deviations from the acceptance criteria detailed in this document will be noted in the **COMMENTS** section at the end of the PQ protocol. All resolution to such problems will also be noted in the **COMMENTS** section, and must be resolved prior to issuance of a **SYSTEM CERTIFICATION**.

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PD), Dist, Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkatoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



AGAPPE

III. SCOPE

This Performance Qualification protocol will be performed on the **MISPA ACE** at **SOM DATT MEDICAL CENTRE, NEW DELHI**. This Performance protocol will define the documentation that will be used to evaluate the instruments installation in accordance with the manufacturer's specifications and intended use. Successful completion of this protocol will verify that the instrument identified has been installed in accordance with the intended usage. Trained, knowledgeable personnel will perform qualification studies. Any exceptional conditions encountered during the qualification studies will be identified for review. Exceptional conditions will be investigated and the appropriate course of action determined. All documents will be initialed and dated.

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. | ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,
 Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
 S-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in



IV. PERFORMANCE QUALIFICATION

A. Instrument Identification

Verified Date : 18/May/2017

1. Model Name : MISPA ACE

B. Serial Number : WK-67104827T

C. Following is a list of tests to be performed and verified:

Test No.	Test Name	Test purpose	Verified By and date
01	Start up	To make the equipment ready for operation	Vijender Pratap Singh 18/May/2017
02	Daily maintenance	To clean appropriate modules	Vijender Pratap Singh 18/May/2017
03	Cell blank	To check the cuvettes	Vijender Pratap Singh 18/May/2017
04	Priming	To check the tubing's, and flow	Vijender Pratap Singh 18/May/2017
05	Auto gain	To check the cuvette and photometry system	Vijender Pratap Singh 18/May/2017
06	Calibration	To calibrate the system	Vijender Pratap Singh 18/May/2017
07	QC check	To confirm that the system and reagents are acceptable and working within specifications	Vijender Pratap Singh 18/May/2017
08	Reproducibility check	To check the precision [CV %]	Vijender Pratap Singh 18/May/2017
09	Sample programming and Analysis	To run the samples	Vijender Pratap Singh 18/May/2017
10	Shut down procedure	To shut down the system	Vijender Pratap Singh 18/May/2017

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. | ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)

401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT

X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 72L. Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE

5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



Test 01: Starting the system

Purpose: To make the instrument READY for operation

Summary:

Instrument checks functioning of different parts of the instrument automatically; if there is an error code, initialize the system and follow corrective action instructions provided for the error message.

Procedure:

- Wait for the instrument to get ready after initialization
- Check the room temperature and switch on the Air Conditioner
- Check the UPS.
- Switch on the MISPA ACE by pressing the main switch, then switch on the computer and then the system power
- Double click the analyzer icon to initialize the system software
- Input the password and do as prompted, the system menu pops up if start up is finished
- If not, initialize by again after solving the error displayed
- Follow instructions provided for the error message

Acceptance criteria: System to display **READY** status

PARAMETER	PASS	FAIL
-----------	------	------

Parameter values for verification: "READY" on Status Area

Joby
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 02: Daily Maintenance**

Purpose: To clean appropriate modules so as per the daily maintenance protocol

Method:

Refer detailed procedure for Daily Maintenance

S No.	Activity	Done by	Date
01	Empty waste container	Vijender Pratap Singh	18/May/2017
02	Check Wash solutions	Vijender Pratap Singh	18/May/2017
03	Check system water	Vijender Pratap Singh	18/May/2017
04	Check Reagents	Vijender Pratap Singh	18/May/2017
05	Clean Reagent Block	Vijender Pratap Singh	18/May/2017
06	Clean out side area	Vijender Pratap Singh	18/May/2017
07	Load supplies and remove outdated and empty reagents	Vijender Pratap Singh	18/May/2017
08	Check main menu screen	Vijender Pratap Singh	18/May/2017
09	Perform Quality Control	Vijender Pratap Singh	18/May/2017

Acceptance criteria: System should be "Ready" after daily maintenance without any error

PARAMETER PASS FAIL

Parameter values for verification: System found "Ready" after daily maintenance

Deep
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PD), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, XINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 03: Cell Blank**

Purpose: To check the cuvette blank OD at different wavelength

Procedure:

S No.	Activity	Done By	Date
01	Checking the Cell blank	Vijender Pratap Singh	18/May/2017
	340 nm	Vijender Pratap Singh	18/May/2017
	405 nm	Vijender Pratap Singh	18/May/2017
	450 nm	Vijender Pratap Singh	18/May/2017
	510 nm	Vijender Pratap Singh	18/May/2017
	546 nm	Vijender Pratap Singh	18/May/2017
	578 nm	Vijender Pratap Singh	18/May/2017
	630 nm	Vijender Pratap Singh	18/May/2017
	670 nm	Vijender Pratap Singh	18/May/2017

Acceptance criteria:

- No error Messages displayed
- All Cuvettes should show acceptance

PARAMETER **PASS** **FAIL**

Parameter values for verification: No Error Messages

Done
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PD), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
K/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 04: Priming**

Purpose: To check the internal tubing's and flow of liquids

S No	Activity	Done By	Date
01	Do wash using the maintenance screen	Vijender Pratap Singh	18/May/2017

Acceptance criteria:

- No error message or air bubbles in the internal tubing's
- All the dilutor working well without air bubbles / leakage

PARAMETER PASS FAIL

Parameter values for verification: No Error Messages

Test 05: Auto gain

Purpose: To Check the photometry system

Procedure:

S No.	Activity	Done By	Date
01	Performing the filter OD checking	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Data shows OD values in the acceptable range

PARAMETER PASS FAIL

Parameter values for verification : No error message

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 06: Calibration of the system****Purpose:** To calibrate the system**Procedure:**

S No.	Activity	Done By	Date
01	Preparation of the cal material *	Vijender Pratap Singh	18/May/2017
02	Performing Calibration with calibration programming screen	Vijender Pratap Singh	18/May/2017

*Traceability of the calibrator used is attached

Acceptance criteria: Calibration data shows concordance OD values in triplicate measurement

PARAMETER PASS FAIL

Parameter values for verification : No error message in the printout

Test 07: QC check**Purpose:** To confirm that systems, reagents and consumables are acceptable & working within specifications for each assay used.**Procedure:**

S No.	Activity	Done By	Date
01	Preparing Biochemistry control material	Vijender Pratap Singh	18/May/2017
02	Creating QC file	Vijender Pratap Singh	18/May/2017
03	QC sample programming and analysis	Vijender Pratap Singh	18/May/2017

Acceptance criteria: QC results within specified limits mentioned on the control product insert

PARAMETER PASS FAIL

Parameter values for verification: QC values within $\pm 2SD$

Pass
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 08: Reproducibility Check****Purpose:** To check the reproducibility of instrument

S No.	Activity	Done By	Date
01	Preparing control material/ sample	Vijender Pratap Singh	18/May/2017
02	Running triplicates	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Results CV within specified limits

PARAMETER	PASS	FAIL
------------------	-------------	-------------

Parameter values for verification: CV% within the limits**Test 09: Sample programming and Analysis****Purpose:** To run the samples**Procedure:**

S No.	Activity	Done By	Date
01	Preparing and Processing of samples	Vijender Pratap Singh	18/May/2017
02	Programming samples	Vijender Pratap Singh	18/May/2017
03	Aspirating the samples	Vijender Pratap Singh	18/May/2017
04	Viewing samples in process	Vijender Pratap Singh	18/May/2017
05	Review results: Monitoring results	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Samples Analysis without any error

PARAMETER	PASS	FAIL
------------------	-------------	-------------

Parameter values for verification: Sample analysis without any error

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119, Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
Tel: +91 33-4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**Test 10: Shut down****Purpose:** To shut down the system**Procedure:**

S No.	Activity	Done By	Date
01	Preparing the system for shut down	Vijender Pratap Singh	18/May/2017
02	Following the on screen instructions	Vijender Pratap Singh	18/May/2017
03	Switch off the instrument when prompted	Vijender Pratap Singh	18/May/2017

Acceptance criteria: Shut down without any error

PARAMETER	PASS	FAIL
-----------	------	------

Parameter values for verification: Shut down without any error

OPERATIONAL PROCEDURE**Certificate of Training****1. Technician Training**

This certifies that the technicians have received basic user training in the following categories for the system described in this Operational Qualification.

S No.	Training program	Initials	Date
1.	Instrument Setup	Vijender Pratap Singh	18/May/2017
2.	System Operation	Vijender Pratap Singh	18/May/2017
3.	Basic trouble shooting and Maintenance	Vijender Pratap Singh	18/May/2017

Jay
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



2. Operator Training

The users responsible for the operation of this equipment have been trained in the proper usage of the system. Training focused on the basic operation and maintenance of the system.

S No.	Operators	Department	Initials	Date

Jays
28/5/22

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in



V. SYSTEM CERTIFICATION

Study data has determined that the system described in this document either meets all criteria outline in this Performance Qualification Protocol, or exceptional conditions have been identified and documentation included. Exceptional conditions, if any, have been addressed. The system is ready for specified usage.

Report Performed By: **AGAPPE DIAGNOSTICS LTD.** Representative

Name : Vijender Pratap Singh

Designation : Engineer - Customer Service

Company : AGAPPE DIAGNOSTICS LTD.

Signature :

Date : 18/May/2017



Customer Authorizations: **SOM DATT MEDICAL CENTRE, NEW DELHI**

Name : *JAYA ULLAS*

Designation :

Title : PERFORMANCE QUALIFICATION

Signature :

Date :

Jays

Jays

28/5/22

Jays

AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/589-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4358 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

Real-time

Daily QC

Day to Day QC

QC Sum.

Control

Test Result

Test Statistics

Date

6/ 1/2022

Control

QC NROM

6/ 1/2022

Test	Actual Result	Unit	Request Date	Ref. Range
SGOT	31.57	U/L	6/1/2022 12:06:40 PM	31.0000-45.0000
SGPT	40.19	U/L	6/1/2022 11:48:04 AM	39.6000-57.6000
GGT	37.51	U/L	6/1/2022 10:59:20 AM	31.5000-45.5000
ALB	3.87	g/dL	6/1/2022 10:59:20 AM	3.4500-4.6500
TPR	5.84	g/dL	6/1/2022 10:59:20 AM	5.4500-7.3500
ALP	144.80	U/L	6/1/2022 10:59:20 AM	126.0000-190.0000
PHOS	3.21	mg/dL	6/1/2022 10:59:20 AM	2.4800-3.9200
BILI-D	0.11	mg/dL	6/1/2022 10:59:20 AM	0.0300-0.2500
BILI-T	1.03	mg/dL	6/1/2022 10:59:20 AM	0.6500-1.2500
CAL	8.25	mg/dL	6/1/2022 10:59:20 AM	7.5800-9.4200
CREAT	0.96	mg/dL	6/1/2022 10:59:20 AM	0.6500-1.1500
URIC ACID	5.28	mg/dL	6/1/2022 10:59:20 AM	4.3000-6.1000
UREA U.V	26.36	mg/dL	6/1/2022 10:59:20 AM	18.2000-33.8000
HDL-D	46.9	mg/dL	6/1/2022 10:59:20 AM	41.700-60.300
TGL	100.3	mg/dL	6/1/2022 10:59:20 AM	85.000-123.000
CHO	172.0	mg/dL	6/1/2022 10:59:20 AM	147.000-193.000
GLU	74.4	mg/dL	6/1/2022 10:59:20 AM	69.000-93.000

View searched QC test record

Refresh

Print

Delete

Reac. Curve

Test Result Date: Control:
 Test Statistics

Test	Actual Result	Unit	Request Date	Ref. Range
GGT	143.87	U/L	6/1/2022 10:15:29 AM	127.0000-177.0000
ALB	4.54	g/dL	6/1/2022 10:15:29 AM	4.1100-5.8900
TPR	8.69	g/dL	6/1/2022 10:15:29 AM	7.9000-11.5000
ALP	429.74	U/L	6/1/2022 10:15:29 AM	372.0000-532.0000
SGPT	146.44	U/L	6/1/2022 10:15:29 AM	131.0000-209.0000
SGOT	185.46	U/L	6/1/2022 10:15:29 AM	177.0000-263.0000
PHOS	8.78	mg/dL	6/1/2022 10:15:29 AM	7.9100-10.1100
BILI-D	0.29	mg/dL	6/1/2022 10:15:29 AM	0.1000-0.6000
BILI-T	4.32	mg/dL	6/1/2022 10:15:29 AM	3.0100-4.9900
CAL	11.73	mg/dL	6/1/2022 10:15:29 AM	10.3000-13.7000
CREAT	2.90	mg/dL	6/1/2022 10:15:29 AM	2.0000-3.6000
URIC ACID	11.59	mg/dL	6/1/2022 10:15:29 AM	10.1500-13.6500
UREA U.V	80.15	mg/dL	6/1/2022 10:15:29 AM	68.0000-96.0000
HDL-D	55.7	mg/dL	6/1/2022 10:15:29 AM	42.720-78.720
TGL	411.4	mg/dL	6/1/2022 10:15:29 AM	373.000-533.000
CHO	216.2	mg/dL	6/1/2022 10:15:29 AM	186.000-254.000
GLU	168.1	mg/dL	6/1/2022 10:15:29 AM	158.700-216.700

View searched QC test record