



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name	<b>SRI SAMRAJ LABS</b>	Lab No	<b>16144</b>
Constituent Group	<b>HbA1c</b>	Date of Result Entered :	<b>19/05/2022</b>
PT item	<b>Lyophilized Whole blood</b>	Date of Report Published :	<b>08/06/2022</b>

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	HPLC-Ion Exchange	Any Analyser	97	6.41	9.77	0.63	6.7 %	0.46	0.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

**SRI SAMRAJ LABS**  
**14/2, NGO COLONY, TINDIVANAM**  
**VILLUPURAM**  
**TAMILNADU604001**

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*