



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name	SRI SAMRAJ LABS	Lab No	16144
Constituent Group	Biochemical Markers for Downs screening	Date of Result Entered :	19/05/2022
PT item	Lyophilized Serum	Date of Report Published :	08/06/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	ALPHA FETOPROTEIN	Other Immunoassays(Any Other Principles / Methods)	Any Analyser	11	50.24	8.76	4.40	46.31 IU/mL	-0.89	2.65
2	FREE BETA HCG	Other Immunoassays(Any Other Principles / Methods)	Any Analyser	21	29.35	8.85	2.60	28.99 ng/mL	-0.14	1.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

SRI SAMRAJ LABS
14/2, NGO COLONY, TINDIVANAM
VILLUPURAM
TAMILNADU604001

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****