



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 3752

Distribution No.: 154-J

Month/Year: January/2022

Instrument ID: XN-550(S.No-20102)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 16-03-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	7.93	7.46	15.39	14.9	0.0430	0.51	0.47	0.11	0.0110	3.74
RBC x10 ⁶ /µl	1	3.53	3.43	6.96	6.82	0.0080	0.73	0.1	0.03	0.0030	1.89
Hb g/dl	1	11	10.8	21.8	21.8	0.0240	0.00	0.2	0.1	0.0080	1.35
HCT%	1	37.8	37	74.8	68.6	0.1510	1.86	0.8	0.3	0.0270	1.69
MCV-fl	1	107.9	107.1	215	200.6	0.3570	1.65	0.8	0.5	0.0370	0.45
MCH-Pg	1	31.5	31.2	62.7	64.1	0.0800	-0.79	0.3	0.2	0.0160	0.34
MCHC-g/dl	1	29.2	29.1	58.3	63.6	0.1440	-1.62	0.1	0.3	0.0220	-0.54
Plt. x10 ³ /µl	1	209	197	406	418	1.39	-0.39	12	6	0.39	1.16
Retic %	2	6	5.9	11.9	6.6	0.17	1.20	0.1	0.3	0.02	-0.67

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=3.00 , Poly=48.00 L=3.00, E=2.00, Mono/Promono=1.00 , B1=1.00 P.M.=, Mye=28.00, Meta=8.00, Other=
RBC Morphology	3	Poly: 45 - 58 , Myelo: 8 - 21, Meta: 6 - 13; Lympho/Promyelo: 1 - 10; Blast/nRBC/Eos/Baso/Mono: 0 - 5
Diagnosis	3	Mild hypochromia
		Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytosis, Hypochromia; Mild: Macrocytosis, Poikilocytosis
		CML
		Chronic Myeloid Leukemia

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 154--J	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	237	237	86.92	85.65	4.64	1.69	8.44	12.66
RBC x10⁶/µl	1	237	237	91.14	91.56	4.22	4.22	4.64	4.22
Hb g/dl	1	237	237	92.83	89.87	3.8	3.8	3.37	6.33
HCT%	1	237	237	90.3	84.81	8.44	7.59	1.26	7.6
MCV-fl	1	237	236	94.92	94.92	3.81	2.97	1.27	2.11
MCH-Pg	1	237	237	89.03	88.19	7.17	6.75	3.8	5.06
MCHC-g/dl	1	237	237	90.72	89.45	8.44	5.06	0.84	5.49
Plt. x10³/µl	1	237	237	89.03	91.56	7.59	4.22	3.38	4.22
ReticCount%	2	237	232	96.12	83.19	1.29	10.78	2.59	6.03
PS Assessment	3	237	217	Satisfactory :90.99%, Borderline Sat. :7.17%, Unsatisfactory :1.84%					

***Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----