

**Result Entry Report**

Lab : 640 / SP HITECH PATHOLOGY LABORATORY

Cycle No : 220104

Sample No : 01

**Analyte Scheme / Group :** Clinical Path - Urine Routine Chemistry

**Analyte :** pH

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	5	5.5	-

**Analyte :** Specific Gravity

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	1.025	1.025	-

**Analyte :** Glucose (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	3+	3+	-

**Analyte :** Protein (Urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	1+	2+	-

**Analyte :** Ketone (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Negative	Negative	-

**Analyte :** Blood (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	1+	2+	-

**Analyte :** Bilirubin (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Negative	Negative	µmol/L

**Analyte :** Urobilinogen (urine)

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Normal	Normal	-

**Result Entry Report**

Lab : 640 / SP HITECH PATHOLOGY LABORATORY

Cycle No : 220104


Sample No : 01

**Analyte : Nitrite (urine)**

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Photoelectric colorimetry	Uriplus 200	06/05/2022	Absent	Absent	-

**Analyte : Urine Pregnancy Test**

Method	Analyzer	Date	Standard Unit Result Val	Result Val	Unit
Immunochemistry	-	06/05/2022	Negative	Negative	-



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070