

MK BEST CALIBRATION SERVICES



एमके बेस्ट केलिब्रेशन सर्विसेस

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CERTIFICATE OF CALIBRATION



CERTIFICATE NO	: MKBN/22/06/0530-001	Date of issue	: 11.06.2022
Date of Calibration	: 11.06.2022	Recom.Due Date	: 10.06.2023

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
Customer Details : M/s: SRI SAMRAJ LABS , (A UNIT OF SRI SAMRAJ HEALTH SERVICES PVT LTD) NO :14/2 , N . G . O COLONY , TINDIVANAM - 604 001 .	Description	CENTRIFUGE	
	Range	0 to 6000 rpm	
	Resolution	1 rpm	
	Make	REMI	
	Model	R - 8C PLUS	
	Identification No	SSLTVM/EQP/CB/001	
	Serial No	ZFHN25200	
	Location	LAB	

S.No	Description	Id.No/SI. No	Traceability	Certificate No	Validity
01	Digital Tachometer	MK/CAL-222 / S3335767	National Standards	CRMTL/01/421103387-A5	20.10.2022

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD & ACCEPTANCE LIMIT	
Temperature	20.4°C	Uncertainty ± °C	1.57
Humidity	52 % RH		

CALIBRATION RESULTS

S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	0	0	0
2	1000	1001	1
3	2000	2001	1
4	3000	3001	1
5	4000	4000	0
6	6000	6001	1

Remarks :

- The Expanded Uncertainty associated with the results are calculated at a confidence level of approximately 95% with a Coverage factor k=2.
- The above UUC was Calibrated at Site.

Calibrated by

S. Chandra Bose
S.Chandra Bose
(Calibration Engineer)

Authorised By

L. Magesh
L.Magesh
(MD/QM)



THE STANDARD INSTRUMENTS USED ARE TRACEABLE TO NABL

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION ISO 9001 : 2015.