

# MK BEST CALIBRATION SERVICES



NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

## CERTIFICATE OF CALIBRATION

URL No	CC334022000010877F	Date of Calibration	11.06.2022	Page No	1 of 1
Certificate No	MKBL/22/06/0669-003	Recom. Due Date	10.06.2023	Date of Receipt	11.06.2022
				Date of Issue	11.06.2022

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION	
M/S. SRI SAMRAJ LABS , ( A UNIT OF SRI SAMRAJ HEALTH SERVICE PVT LTD ) NO : 14/2 , N.G.O COLONY , TINDIVANAM - 604 001 .	Description	STEM THERMOMETER
	Range / Resolution	-50 - 300°C / 0.1°C
	Make / Model	MULTI
	Identification No	STM- 01
	Serial No	---
	Identification No	SSLTVM/EQP/CB/004
	Location	CLINIC BIO CHEMISTRY
	Calibrated At	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	SSPRT	MK CAL-193	CRM TL/01/421100630-A1	18.11.2022
02	DIGITAL MULTIMETER	MK/CAL-45	EQN/CRF/2201081	23.01.2023

ENVIRONMENTAL & DUC CONDITIONS		REFERENCE STANDARD	
Temperature	25 ± 4°C	Reference Std	EURAMET/cg-11/v.01
Humidity	30 - 75 % RH	Procedure No	MKBCS - ET - 01
Condition of DUC	Good		

### CALIBRATION RESULTS

#### 1.ELECTRO TECHNICAL CALIBRATION - TEMPERATURE SIMULATION

S.No	Standard Reading °C	DUC Reading (Mean) °C	Deviation °C	Expanded Uncertainty (±) °C
1	-40.00	-39.9	0.1	0.71
2	0.00	0.2	0.2	0.71
3	50.00	50.2	0.2	0.71
4	100.00	100.3	0.3	0.71
5	200.00	200.3	0.3	0.71
6	250.00	250.3	0.3	0.71

#### Remarks :

- The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval Should be Determined on the User Requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The User Should Determine The Suitability Of The Instrument For Its Intended Use.

Calibrated by

S.Chandra Bose  
(Calibration Engineer)

x-x-x-x- End Of Certificate -x-x-x-x



Authorised By

L.Magesh  
(MD/QM)