



CYCLE 52
ROUND: 2

Lab Code No. 2135

BASIC SEROLOGY
FINAL RESULT ASSESSMENT

CYCLE NO. : 11

ROUND: 2

TOTAL PARTICIPANT: 328

DATE: 5/04/2022

Parameter	Total Responses	Your Result	All Lab Result	%	Remarks
S1- CRP	260	Non-Reactive	Reactive : 25 Non-Reactive : 235	90%	Within Consensus
S2- RPR	269	Non-Reactive	Reactive : 43 Non-Reactive : 226	84%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

End of Report

Prepared by: JP

Page 1 of 1

