



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - APRIL 2022



Lab No 6798

Lab Name SRL LIMITED

Date of Result Entered : 06/04/2022

Constituent Group HbA1c

Date of Report Published : 05/05/2022

PT item Lyophilized Whole blood

SI/No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants	SD	CV	CV %	SDI	U
1	HbA1c	OTHERS ( Any Other Principles / Methods )	Any Analyser (Automation / Semi Automation )	227	6.79	12.31	0.84	7.34 %	0.66	0.11		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

SRL LIMITED  
 CEDMAR APARTMENTS, BLOCK D-A , NEXT TO HOTEL ARCADIA, M.G.ROAD  
 PANJIM  
 GOA403001

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*