

LOTUS DIAGNOSTICS (562)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**LOTUS DIAGNOSTICS (562)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 01**


**Report Date : 23/05/2022**

**Instrument : Nihon Kohden Celltac Alpha**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.08	4.93	1.04	0.20
@ Total Leucocytes (WBC) count	/c.mm	5200.00	4,314.00	1.85	0.49
@ Haemoglobin	g/dL	16.30	15.68	1.35	0.53
@ MCV (Mean Corpuscular Volume)	fL	96.30	101.61	-0.64	-0.95
@ MCH (Mean Corpuscular Hb)	pg	32.10	31.69	0.13	0.40
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	33.30	30.90	0.76	1.10
@ Platelet Count	1000/ $\mu$ L	103.00	105.95	-0.94	-0.15
@ Hematocrit	%	48.90	50.79	-0.30	-0.76
@ RDW CV%	%	12.90	12.88	-0.40	-0.67

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**

**PT coordinator & Technical Manager, MHL EQAS**

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report

LOTUS DIAGNOSTICS (562)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210512 and Sample No 05

Report Date : 30/12/2021

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**LOTUS DIAGNOSTICS (562)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210512 and Sample No 05**


**Report Date : 30/12/2021**

**Instrument : -**

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	<40%	<40%
@ Lymphocyte.	-	>40%	>40%
@ Eosinophil.	-	<6%	<6%
@ Monocyte.	-	1-10%	1-10%

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	4
Not Evaluated Parameters	0
Evaluated Parameters	4
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report