

CALIBRATION CERTIFICATE

User Name: M/s Sanjivani Diagnostics
 Shop-2085, Ganjmer Khan, Near Delite Cinema,
 Opposite Asaf Ali Road, Delhi-110002.



Page No.: 1 of 1

Unit Under Calibration	Refrigerator Thermometer	Make / Model No.	---- / ----
ULR NO.	CC205322000000615F	Range/ Size	-10 ~ -20 °C
Certificate No.	KRC/2022/EH/005-002	Least Count	0.1 °C
Service Request Date	18-May-2022	UUC Serial No.	----
Calibration Date	19-May-2022	UUC I.D. No.	RT-02
Suggested Due Date	18-May-2023	Location	----
Date Of Issue	20-May-2022	Visual Inspection	OK

Relevant Standard	IEC 60601-2-21:2009	Calibration Procedure	KRC/CPM/7.2-18-H-WI	Performed At	Site
-------------------	---------------------	-----------------------	---------------------	--------------	------

ENVIRONMENTAL CONDITIONS

Temperature	25 ± 4 °C	Humidity	50 ± 20 % RH
-------------	-----------	----------	--------------

MASTER INSTRUMENT USED FOR CALIBRATION

Instrument Name	Make / Model No.	Serial / I.D. No.	Traceability	ULR No.	Suggested Due Date
PRT with Indicator	Fluke / 5609	05151	FARELABS	FL/C/TH/23112021-020	25-Nov-2022

CALIBRATION RESULTS IN °C


Master Average Reading (X)	UUC Average Reading (Y)	Error = Y-X
-10.021	-10.0	0.021
-12.037	-12.0	0.037
-14.042	-14.0	0.042
-16.055	-16.0	0.055
-18.068	-18.0	0.068
-20.070	-20.0	0.070

Expanded Uncertainty ± 0.06 °C

Uncertainty of Measurement: The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%.

Note:-

- The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
- The results reported relate only to the above calibrated item.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid.
- Rdg. and UUC stands for reading and unit under calibration respectively.

Calibrated By

 Calibration Engineer



END OF REPORT