



# Avi Scientific (India)

CALIBRATION LABORATORY



## CALIBRATION CERTIFICATE

AVI/Certificate/F-12D		Calibration Certificate for Micropipette			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue	ULR No :	CC264521000000202F	
16.07.2021		15.07.2022	17.07.2021	Certificate No.:	AVI2107058	
SRF No.	2100084	SRF Date	16.07.2021	Job No	2107058	Job Date
						16.07.2021
<b>Discipline : Mechanical calibration</b>			<b>Group : Mass &amp; Volume</b>			
Customer's Name & Address:				Lotus Diagnostics shop no.5 Chandresh CHS, near lotus hospital, devki nagar, Borivali west, Mumbai 103		
Calibration Carried out:				In-House		
Description of item Calibrated:						
Make				Thermoscientific		
Model				FINNPIPETTE F3		
Serial No.				MW20290		
Identification No.				LP/MP/01		
Range				100 ( µl )		
Location				---		
Work Instruction No.				AVI/WI/12		
Environmental Condition				Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030		
Standard(s) used for Calibration:				Digital Weighing Balance, Sr No. B125154050 , Cert No.- CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.		

### OBSERVATION TABLE

#### Volume at 27°C

Sl. No.	Range	UUC Reading	Standard Reading	Error	Expanded Uncertainty (At 95% CL, K=2), (±µl)
		( µl )	( µl )	( µl )	
1	100	100	100.31	-0.31	0.50

Calibrated By

(Prathamesh Mestri)

Calibration Engineer



Approved By

( Nandkumar Patil )

Quality Manager

Please Refer Note Backside of the Page