



Avi Scientific (India)

CALIBRATION LABORATORY



CALIBRATION CERTIFICATE

AVI/Certificate/F-12D		Calibration Certificate for Micropipette			Page No: 01 of 01	
Date of Calibration:		Next Cal. due on:	Date of Issue	ULR No :	CC26452100000204F	
16.07.2021		15.07.2022	17.07.2021	Certificate No.:	AVI2107060	
SRF No.	2100084	SRF Date	16.07.2021	Job No	2107060	Job Date 16.07.2021
Discipline : Mechanical calibration			Group : Mass & Volume			
Customer's Name & Address:				Lotus Diagnostics shop no.5 Chandresh CHS, near lotus hospital, devki nagar, Borivali west, Mumbai 103		
Calibration Carried out:				In-House		
Description of item Calibrated:						
Make				Thermoscientific		
Model				FINNPIPETTE F3		
Serial No.				NW04749		
Identification No.				LD/MP/03		
Range				500(µl)		
Location				---		
Work Instruction No.				AVI/WI/12		
Envimmental Condition				Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030		
Standard(s) used for Calibration:				Digital Weighing Balance, Sr No. B125154050 , Cert No.- CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.		

OBSERVATION TABLE

Volume at 27°C

Sl. No.	Range	UUC Reading	Standard Reading	Error	Expanded Uncertainty (At 95% CL, K=2), (±µl)
		(µl)	(µl)	(µl)	
1	500(µl)	500	500.34	-0.34	0.50

Calibrated By

(Prathamesh Mestry)

Calibration Engineer



Approved By

(Nandkumar Patil)

Quality Manager

Please Refer Note Backside of the Page