



Avi Scientific (India)

CALIBRATION LABORATORY



CALIBRATION CERTIFICATE

AVI/Certificate/F-12D	Calibration Certificate for Micropipette			Page No: 01 of 01
Date of Calibration:	Next Cal due on:	Date of Issue	ULR No :	CC264521000000202F
16.07.2021	15.07.2022	17.07.2021	Certificate No.:	AVI2107058
SRF No. 2100084	SRF Date	16.07.2021	Job No	2107058
Discipline : Mechanical calibration		Group : Mass & Volume		
Customer's Name & Address:		Lotus Diagnostics shop no.5 Chandresh CHS, near lotus hospital, devki nagar, Borivali west, Mumbai 103		
Calibration Carried out:		In-House		
Description of item Calibrated:				
Make	Thermoscientific			
Model	FINNPIPETTE F3			
Serial No.	MW20290			
Identification No.	LD/MP/01			
Range	100 (µl)			
Location	---			
Work Instruction No.	AVI/WI/12			
Envimmental Condition	Temp. (°C): (25 ± 2) Relative Humidity (% RH) (40 - 60) Atmospheric Pressure (hpa) 970-1030			
Standard(s) used for Calibration:	Digital Weighing Balance, Sr No. B125154050 Cert No. - CC/M&V/0699/20-21, CC 2287, Due date :15/09/2021 ,calibratd by IDEMI , Mumbai.			

OBSERVATION TABLE

Volume at 27 °C

Sl. No.	Range	UUC Reading	Standard Reading	Error	Expanded Uncertainty (At 95% CL, K=2), (±µl)
		(µl)	(µl)	(µl)	
1	100	100	100.31	-0.31	0.50

Calibrated By *Prathamesh Mestri*
(Prathamesh Mestri)
 Calibration Engineer



Approved By *(Janardan Chavan)*
 (Janardan Chavan)
 Technical Manager

Please Refer Note Backside of the Page