



Installation Qualification

| | |
|-------------------------------------|--|
| Instrument Type: Blood Cell Counter | Manufacture Name: ERMA Japan |
| Model Name: PCE 525 | Supplier Name: AGD Biomedicals (P) LTD |

Operational Verification Report

| | |
|-------------------------------------|--|
| Instrument Type: Blood Cell Counter | Name of the Customer: Shreeji Pathological Laboratory, Malad |
| Serial Number: 610620-000031 | Model Number: PCE 525 |
| Date of Installation: 29.01.2020 | Date of receipt: 28.01.2020 |

Power Check:

| | |
|---------------|----------|
| Power Supply: | 220 V AC |
| Earth Voltage | OK |

Physical Check of the equipment

| | | |
|--------------------------------|---------|-----------------|
| Check for damage | Checked | No damage found |
| Visual & functional inspection | Checked | OK |
| No damage on internal assembly | Checked | No damage found |

System configuration equipment

| | | |
|---|---------|-----------------|
| Positioning of the equipment -- flat stable surface | Checked | Found OK |
| Adequate space around the equipment | Checked | Found OK |
| Interference to other equipment | Checked | No Interference |

Result:

Installed the PCE 525 satisfactorily and is ready for operational verification

| | |
|------------------------------|-------------|
| Installed by: Mr Ajay Sharma | Checked by: |
| Signature: | Signature: |
| Date: 29/01/2020 | Date: |



Operational Qualification

| | |
|-------------------------------------|--|
| Instrument Type: Blood Cell Counter | Manufacture Name: ERMA Japan |
| Model Name: PCE 525 | Supplier Name: AGD Biomedicals (P) LTD |
| Installed On: 25.02.2022 | Service Engineer: Mr Ajay Sharma |

Operational Verification Report

| | |
|--|-------------------------------------|
| Name of the Customer: Shreeji Pathological Laboratory, Malad | Instrument Type: Blood Cell Counter |
| Model Number: PCE 2525 | Serial Number: 610620-000031 |
| Date of receipt: 28.01.2020 | Date of Installation: 29.01.2020 |

Operational Check:

| | |
|-----------------------|-----------------------------|
| Background Check | OK |
| Control Check | All Values are within Range |
| Reproducibility Check | OK |

Result:

Operational verification of PCE 525 has been completed satisfactorily and it is ready for the performance verification

| | |
|--|-------------|
| Operations performed by: Mr. Ajay Sharma | Checked by: |
| Signature: | Signature: |
| Date: 25/2/2022 | Date: |