



Installation Qualification

Instrument Type: New Generation Clinical Chemistry Analyzer	Manufacture Name: AGD Biomedicals (P) LTD
Model Name: AGD2020	Supplier Name: AGD Biomedicals (P) LTD

Operational Verification Report

Instrument Type: New Generation Clinical Chemistry Analyzer	Name of the Customer: Shreeji Pathological Laboratory, Malad
Serial Number: AO1705104	Model Number: AGD2020
Date of Installation: 24.06.2017	Date of receipt: 23.06.2017

Power Check:

Power Supply:	220 V AC
Earth Voltage	OK

Physical Check of the equipment

Check for damage	Checked	No damage found
Visual & functional inspection	Checked	OK
No damage on internal assembly	Checked	No damage found

System configuration equipment

Positioning of the equipment - flat stable surface	Checked	Found OK
Adequate space around the equipment	Checked	Found OK
Interference to other equipment	Checked	No Interference

Result:

Installed the AGD 2020 satisfactorily and is ready for operational verification

Installed by: Mr Ajay Sharma	Checked by:
Signature:	Signature:
Date: 24/6/2017	Date:



Operational Qualification

Instrument Type: New Generation Clinical Chemistry Analyzer	Manufacture Name: AGD Biomedicals (P) LTD
Model Name: AGD2020	Supplier Name: AGD Biomedicals (P) LTD
Installed On: 24.06.2017	Service Engineer: Mr Ajay Sharma

Operational Verification Report

Name of the Customer: Shreeji Pathological Laboratory, Malad	Instrument Type: New Generation Clinical Chemistry Analyzer
Model Number: AGD2020	Serial Number: AO1705104
Date of receipt: 23.06.2017	Date of Installation: 24.06.2017

Operational Check:

Reproducibility Check	OK
Standard Results	OK

Result:

Operational verification of AGD2020 has been completed satisfactorily and it is ready for the performance verification

Operations performed by: Mr. Ajay Sharma	Checked by:
Signature:	Signature:
Date: 24/6/2017	Date: