

Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,ASO Qualitative, IU/mL

	All Methods	N
	Agglutination - (Your Method)	69
	Manual - (Your Analyzer)	38
		39

Your Result	Normal
Result Expected	Nor

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-
 Nor : Normal
 Abn : Abnormal

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022


Immunoassay Extended,C-Reactive Protein Qualitative, mg/L

	N
█ All Methods	120
█ Agglutination - (Your Method)	25
█ Manual - (Your Analyzer)	31

Evaluation Based On - Equipment wise Peer Grouping	
Your Result	Normal
Result Expected	#

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-
 Nor : Normal
 Abn : Abnormal

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,Dengue NS 1, -

	N
■ All Methods	156
■ Immunochromatography - (Your Method)	137
■ Manual - (Your Analyzer)	148

Your Result	Negative
Result Expected	Neg

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

- Pos : Positive
- Eq : Equivocal
- BL : Borderline
- Neg : Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,HBsAg, -

	All Methods	N 204
	Immunochromatography - (Your Method)	150
	Manual - (Your Analyzer)	133

Your Result	Non-reactive
Result Expected	NR

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

R : Reactive
 NR : Non-reactive
 Eq : Equivocal

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,HCV, -

	N
■ All Methods	154
■ Immunochromatography - (Your Method)	107
■ Manual - (Your Analyzer)	96

Your Result	Non-reactive
Result Expected	NR

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

R : Reactive
 NR : Non-reactive
 Eq : Equivocal

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,HIV Ab, -

	All Methods	N 118
	Immunochromatography - (Your Method)	110
	Manual - (Your Analyzer)	102

Your Result	Non-reactive
Result Expected	NR

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

R	: Reactive
NR	: Non-reactive
Eq	: Equivocal
Indet	: Indeterminate
HIV-1	: HIV-1 Ab Reactive
HIV-2	: HIV-2 Ab Reactive

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,HIV Ag + Ab, Nil

	All Methods	N
	Immunochromatography - (Your Method)	91
	Manual - (Your Analyzer)	37
		36

Your Result	Non-reactive
Result Expected	NR

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022



Reported Results Legends :-

R : Reactive
 NR : Non-reactive
 Eq : Equivocal
 Indet : Indeterminate

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,Rheumatoid Factor Qualitative, cutoff index

	N
■ All Methods	103
■ Immunoturbidimetric - (Your Method)	28
■ Manual - (Your Analyzer)	49

Your Result	Normal
Result Expected	Nor

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

Eq : Equivocal
Nor : Normal
Abn : Abnormal

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 1
 Sample Date : 26-04-2022


Immunoassay Extended,RPR (Syphilis Serology), -

	N
■ All Methods	172
■ Immunochromatography - (Your Method)	65
■ Manual - (Your Analyzer)	172

Your Result	Non-reactive
Result Expected	NR

Note : # is an indication that this parameter has not been evaluated


Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 1
Sample Date : 26-04-2022



Reported Results Legends :-

R : Reactive
NR : Non-reactive
WR : Weak Reactive

Comments :

- Interpretation when an analyte is reported for multiple samples in a Round:
- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
 - b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance. Review for systematic error

Dr. Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070



This is a computer generated report hence signature is not required

Analyte Methodwise Qualitative Result Report

Lab No.: 650
Lab Name : DR.KAUSHIK PATH LABS
Report Date : 23/05/2022

Cycle No : 220104
Sample No : 2
Sample Date : 26-04-2022


Immunoassay Extended,Typhidot IgM, -

	N
■ All Methods	161
■ Immunochromatography - (Your Method)	158
■ Manual - (Your Analyzer)	161

Your Result	Negative
Result Expected	Neg

Note : # is an indication that this parameter has not been evaluated



Frequency Histogram



Analyte Methodwise Qualitative Result Report

Lab No.: 650
 Lab Name : DR.KAUSHIK PATH LABS
 Report Date : 23/05/2022

Cycle No : 220104
 Sample No : 2
 Sample Date : 26-04-2022



Reported Results Legends :-
 Pos : Positive
 Neg : Negative

Comments :

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
 - Two out of two pass - Acceptable performance
 - One out of two pass - Review for random error & take action if appropriate
 - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
 - All pass - Acceptable performance
 - Two out of three pass - Review for random error & take action if appropriate
 - One out of three pass - Unacceptable performance, Review for systematic error
 - All failed - Unacceptable performance, Review for systematic error

AK Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070



This is a computer generated report hence signature is not required