

Date - 20/06/22

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : UREA

d. Abs : 0.0564

Result: 34.53U/L

+1SD

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : CREAT

d. Abs : 0.0564

Result: 1.73U/L

-2SD

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : GLUCOSE

Result: 89.2mg/dL

+1SD

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : CHOL

Result: 266.45mg/dL

+2SD

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : TRIG

Result: 181.41mg/dL

-1SD

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : SGOT

dA/Min: 0.0564

Result: 38.63U/L

-1SD

MERILYZER

CliniQuant QC Report

Date : 20/06/22

Level : M

Test : SGPT

dA/Min: 0.0564

Result: 28.65U/L

-1SD

Date - 21/06/22

MERILYZER

CliniQuant QC Report

Date : 21/06/22

Level : M

Test : UREA

d. Abs : 0.3859

Result: 33.29U/L

+1SD

MERILYZER

CliniQuant QC Report

Date : 21/06/22

Level : M

Test : CREAT

d. Abs : 0.3859

Result: 1.76U/L

-2SD

Date - 09/06/22
Date - 13/06/22

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : UREA
d. Abs : 0.1277
Result: 35.83U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : CREAT
d. Abs : 0.1277
Result: 1.82U/L
-1SD

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : GLUCOSE
Result: 96.14mg/dl
+1SD

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : CHOL
Result: 255.29mg/dL
+1SD

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : TRIG
Result: 160.39mg/dL
-2SD

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : SGOT
dA/Min: 0.1277
Result: 39.84U/L
-1SD

MERILYZER
CliniQuant QC Report
Date : 13/06/22
Level : M
Test : SGPT
dA/Min: 0.1277
Result: 36.06U/L
+2SD

Date - 14/06/22

MERILYZER
CliniQuant QC Report
Date : 14/06/22
Level : M
Test : UREA
d. Abs : 0.1277
Result: 35.72U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 14/06/22
Level : M
Test : CREAT
d. Abs : 0.1277
Result: 1.81U/L
-1SD

Date - 15/06/22

MERILYZER
CliniQuant QC Report
Date : 14/06/22
Level : M
Test : UREA
d. Abs : 0.1864
Result: 36.87U/L
+2SD

MERILYZER
CliniQuant QC Report
Date : 14/06/22
Level : M
Test : CREAT
d. Abs : 0.0882
Result: 1.81U/L
-1SD

Date - 09/06/22

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : UREA
d. Abs : 0.0919
Result: 29.6U/L
-2SD

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : CREAT
d. Abs : 0.0919
Result: 1.82U/L
-1SD

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : GLUCOSE
Result: 97.27mg/dL
+2SD

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : SGOT
dA/Min: 0.0919
Result: 43.72U/L
+1SD

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : SGPT
dA/Min: 0.0919
Result: 31.77U/L
+1SD

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : CHOL
Result: 253.36mg/dL
+1SD

MERILYZER

CliniQuant QC Report

Date : 09/05/22
Level : M
Test : TRIG
Result: 161.81mg/dL
-2SD

Date - 10/06/22

MERILYZER

CliniQuant QC Report

Date : 10/06/22
Level : M
Test : CREAT
d. Abs : 0.0408
Result: 1.77U/L
-2SD

MERILYZER

CliniQuant QC Report

Date : 10/06/22
Level : M
Test : UREA
d. Abs : 0.0408
Result: 32.34U/L
-1SD

Date - 11/06/22

MERILYZER

CliniQuant QC Report

Date : 11/05/22
Level : M
Test : UREA
d. Abs : 0.0467
Result: 28.3U/L
-2SD

MERILYZER

CliniQuant QC Report

Date : 11/05/22
Level : M
Test : CREAT
d. Abs : 0.0467
Result: 2U/L
-1SD

Date 30/05/22

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : UREA
d.Abs : 0.1537
Result: 35.31U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : CREAT
d.Abs : 0.0919
Result: 1.61U/L
-2SD

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : SGOT
dA/Min: -0.0196
Result: 42.01U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : SGPT
dA/Min: -0.0153
Result: 30.88U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : CHOL
Result: 236.1mg/dL
-1SD

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : TRIG
Result: 169.61mg/dL
-1SD

MERILYZER
CliniQuant QC Report
Date : 30/05/22
Level : M
Test : GLUCOSE
Result: 101.44mg/dL
+2SD

Date - 31/05/22

MERILYZER
CliniQuant QC Report
Date : 31/05/22
Level : M
Test : UREA
d.Abs : 0.0919
Result: 32.76U/L
-1SD

MERILYZER
CliniQuant QC Report
Date : 31/05/22
Level : M
Test : CREAT
d.Abs : 0.0919
Result: 1.53U/L
-2SD

Date - 01/06/22

MERILYZER
CliniQuant QC Report
Date : 01/06/22
Level : M
Test : UREA
d.Abs : 0.056
Result: 36.9U/L
+2SD

MERILYZER
CliniQuant QC Report
Date : 01/06/22
Level : M
Test : CREAT
d.Abs : 0.035
Result: 1.53U/L
-2SD

Date - 12/05/22

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : CREAT
d. Abs : 0
Result: 1.57U/L
-2SD

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : UREA
d. Abs : 0
Result: 39.08U/L
+2SD

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : TRIG
Result: 176.37mg/dL
-1SD

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : SGPT
dA/Min: 0
Result: 33.43U/L
+1SD

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : SGOT
dA/Min: 0
Result: 45.22U/L
+1SD

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : CHOL
Result: 233.45mg/dL
-1SD

MERILYZER

CliniQuant QC Report

Date : 12/05/22
Level : M
Test : GLUCOSE
Result: 97.23mg/dL
+2SD

Date 13/05/22

MERILYZER

CliniQuant QC Report

Date : 13/05/22
Level : M
Test : UREA
d. Abs : 0.2035
Result: 34.27U/L
+1SD

MERILYZER

CliniQuant QC Report

Date : 13/05/22
Level : M
Test : CREAT
d. Abs : 0.1992
Result: 1.56U/L
-2SD

Date - 14/05/22

MERILYZER

CliniQuant QC Report

Date : 14/05/22
Level : M
Test : UREA
d. Abs : -0.0132
Result: 36.72U/L
+2SD

MERILYZER

CliniQuant QC Report

Date : 14/05/22
Level : M
Test : CREAT
d. Abs : -0.0117
Result: 1.97U/L
-1SD

Date ⇒ 05/05/2022

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : CREAT
d. Abs : 0.0388
Result : 1.49U/L
-2SD

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : UREA
d. Abs : 0.0719
Result : 35.41U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : GLUCOSE
Result : 86.9mg/dL
-1SD

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : SGOT
dA/Min : 0.0388
Result : 46.47U/L
+1SD

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : SGPT
dA/Min : 0.0388
Result : 25.22U/L
-1SD

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : TRIG
Result : 195.23mg/dL
+1SD

MERILYZER
CliniQuant QC Report
Date : 05/05/22
Level : M
Test : CHOL
Result : 221.24mg/dL
-2SD

Date ⇒ 06/05/2022

MERILYZER
CliniQuant QC Report
Date : 06/05/22
Level : M
Test : UREA
d. Abs : 0.0543
Result : 39U/L
+2SD

MERILYZER
CliniQuant QC Report
Date : 06/05/22
Level : M
Test : CREAT
d. Abs : 0.0388
Result : 1.47U/L
-2SD

Date ⇒ 07/05/2022

MERILYZER
CliniQuant QC Report
Date : 07/05/22
Level : M
Test : UREA
d. Abs : 0.0388
Result : 37.72U/L
+2SD

MERILYZER
CliniQuant QC Report
Date : 07/05/22
Level : M
Test : CREAT
d. Abs : -0.0239
Result : 1.49U/L
-2SD