



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name **ESIC MEDICAL COLLEGE AND HOSPITAL**

Lab No **7760**

Constituent Group **Chemistry I**

Date of Result Entered : **15/05/2022**

PT item **Lyophilized Serum**

Date of Report Published : **08/06/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	296	179.51	4.12	7.39	160 mg/dL	-2.64	0.86
2	UREA	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	303	64.05	4.50	2.88	60 mg/dL	-1.41	0.33
3	CREATININE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	298	2.49	4.61	0.12	2.4 mg/dL	-0.78	0.01
4	T.BILIRUBIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	297	3.23	8.12	0.26	2.86 mg/dL	-1.41	0.03
5	T-PROTEIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	295	5.54	4.59	0.25	5.6 g/dL	0.24	0.03
6	ALBUMIN	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	295	3.32	4.89	0.16	3.2 g/dL	-0.74	0.02
7	CALCIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	292	10.85	3.81	0.41	10.6 mg/dL	-0.61	0.05
8	PHOSPHORUS	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	245	4.83	5.76	0.28	5.2 mg/dL	1.33	0.04
9	URIC ACID	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	291	4.68	3.59	0.17	4.6 mg/dL	-0.48	0.02
10	CHOLESTEROL	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	275	115.56	5.15	5.95	122 mg/dL	1.08	0.72
11	TRIGLYCERIDE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	275	124.13	5.18	6.43	144 mg/dL	3.09	0.78
12	HDL CHO	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	271	23.24	6.71	1.56	23 mg/dL	-0.15	0.19
13	SODIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	228	139.36	2.87	4.00	145 mmol/L	1.41	0.53
14	POTASSIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	230	4.82	2.86	0.14	5 mmol/L	1.30	0.02
15	CHLORIDE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	177	100.24	2.68	2.68	103 mmol/L	1.03	0.40
16	AST	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	296	90.21	6.78	6.12	93 U/L	0.46	0.71
17	ALT	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	295	57.87	9.95	5.76	46 U/L	-2.06	0.67

18	ALP	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	300	91.80	8.27	7.60	81 U/L	-1.42	0.88
19	AMYLASE	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	163	47.62	12.67	6.03	51 U/L	0.56	0.95
20	IRON	Dry Chemistry I	Ortho Clinical Diagnostics / Fuji - Dry Chemistry Series	85	89.78	9.63	8.64	79 ug/dL	-1.25	1.87
21	MAGNESIUM	Dry Chemistry	Ortho Clinical Diagnostics Dry Chemistry Series	112	2.63	5.36	0.14	2.5 mg/dL	-0.92	0.03

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

ESIC MEDICAL COLLEGE AND HOSPITAL
 DR. HARNAM KAUR, DEPAT OF BICHEMISTRY
 FARIDABAD
 HARYANA121001

Coordinator Contact Details:
 Email: clinq@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****