

150/MI/2022-23/368

May 21, 2022

The Dean
ESIC Medical College and Hospital
NH-3, NIT
Faridabad – 121 001.



CERTIFICATE OF CALIBRATION

Name and Address of Customer : ESIC Medical College and Hospital, Faridabad
Location : Lab
Model : Automated Hematology Analyzer Sysmex XP100
Make : Sysmex, Japan
Serial No : A9639
Calibration Date : 21.05.2022
Next Calibration Date : 20.05.2023
Material used : SCS1000 (Lot No. 21090525)
Expiry Date : 22.05.2022

Note: This is to certify that the above mentioned Instrument installed in the lab has been calibrated using the above mentioned Calibrator satisfactorily. The following procedures were done and passed ok.

Sr No.	Procedure	Pass
1	Background Check	OK
2	Precision	OK
3	Calibrator Run	OK
4	Calibration Check 5 times	OK
5	Tri Level QC run	OK

For Medex India Pvt Ltd


A S S D

GST No.: 07AAACM9017E1ZP ★ ROC No.: U51507DL1995PTC071531 ★ Drug Licence No.: DL-TGB-149920

Registered Office : DD-36, Kalkaji, New Delhi-110019(INDIA)

Branches : ★ BANGALORE ★ CHANDIGARH ★ HAMIRPUR (H.P.) ★ KOLKATA ★ ODISHA (GST No. for Odisha 21AAACM9017E1ZZ)