

Accurex Biomedical Pvt. Ltd.

212, Udyog Mandir No 1, 7/C, Bhagoji Keer Marg, Mahim, Mumbai-400016.
Ph : +91 (022) 6744 6744, Fax : +91 (022) 6744 6755



Call Attendance Report

Job No.		Report No.	6129
Date	06/08/2021	Name of Engineer	MR AKSHAY D MENGE
Type of Customer	Warranty	AMC	Non AMC

Laboratory Name & Address		Complaint Received	Purpose of Visit				
SRL DIAGNOSTICS M.G. Road, Near woodland showroom, Panaji, Goa-403001		d d m m y y	<input checked="" type="checkbox"/>	Installation			
		Call attended on		<input checked="" type="checkbox"/>	Test		
		d d m m y y	<input checked="" type="checkbox"/>	Emergency Call			
		Duration h h m m	<input checked="" type="checkbox"/>	Pre-Installation			
		d d m m y y	<input checked="" type="checkbox"/>	PM	1st	2nd	3rd
Contact person	Reema Verlekar	Duration	h	h	m	m	
Contact No.	9545785012	Work Carried Out At		Others (specify)			
Email ID	reema.verlekar@sr-in	<input checked="" type="checkbox"/>	Site	<input checked="" type="checkbox"/>	Office		
Instrument Model	SPPERA	Serial No.		201457002			
Nature of Complaint	Installation - old machine	Any damage noticed before service		No			
Engineer's diagnosis	-						
Action taken	Connected all supply & water supply. machine Initialization ok. Startup ok. Calibration done. Run Q&N & Q-C-H results on target.						

Part Replacement & Repair Estimate Approval

S/N	Code No	Parts Description	Qty	Unit Price (Rs.)	Total Price (Rs.)
 					
 					
 					

The above mentioned part replacement & total service charge have been approved to proceed for necessary service/repair.

Cost of Parts	
Service Charge	
Tax	
Total	


Name		Signature	
------	--	-----------	--

FOR BILLABLE TRANSACTIONS : Customer may kindly note that this CAR duly signed and stamped will be treated as Purchase Order from you. If it does not meet your internal procedural requirement, you may treat this as quotation and issue us a separate order. This estimate is valid for next 30 days from date mentioned here.

Status after the Service	<input checked="" type="checkbox"/>	Functioning Normally	<input checked="" type="checkbox"/>	Functioning but requires action
	<input checked="" type="checkbox"/>	Under Observation	<input checked="" type="checkbox"/>	Referred to Service Centre
Pending job, if any	-			

To avail continued engineering services of Accurex AMC / RR to be renewed from

for Accurex Biomedical Pvt. Ltd.



Remarks		Remarks	
Signature	<i>[Signature]</i>	Signature & Seal	
Name	MR AKSHAY D. MENGE	Name	
Date	06/08/2021	Date	d d m m y y y y

Customer Copy