

**Calibration Certificate**

ADL/BO/MS/C/0297/21-22

This is to certify that **Mispa-i2 Specific Protein Analyzer** with Serial No:
21161631141 at **SRL Limited, Panjim** have been inspected and calibrated.

Calibration Details

Instrument Name	: Mispa-i2 Specific Protein Analyzer
Installation Date	: 15/02/2016
Calibration Date	: 04/Feb/2022
Calibration Certificate Due Date	: 03/Aug/2022
Checked the IC Card Reader	: Found OK
Checked Electronic Pipette Jack	: Found OK
Checked the Cuvette Motor	: Found OK
Checked Mother Board	: Found OK
Checked with Control	: Found OK

For Agappe Diagnostic Ltd.

Authorized Signatory



Name: *Bibin Baby*

Designation: *Sr. Engineer - Customer Service*

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT

Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)

401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sabar Road, Parshwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbai@agappe.in

EQUIPMENT PLANT

X/586-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721, Tel: +91 484 276 7477.

DELHI OFFICE

DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhi@agappe.in

KOLKATA OFFICE

406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091,
Tel: +91 33 4003 0451 | Email: kolkata@agappe.in

BANGALORE OFFICE

5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001, Tel: +91 80 2228 8288
Email: bangalore@agappe.in