

TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C

A NABL Accredited Laboratory

Tel. : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Calibration Certificate

Calibration field: Thermal-Temperature	NABL Accreditation no: CC-2112
Accreditation valid upto: 11.04.2024	Certificate no TR/2022/TEMP/297
Requisition no: Supply	ULR NO: CC211222000001646
Calibration done at At Lab	Date of calibration: 03.06.2022
Date of Receipt 03.06.2022	Agreed Next Calib On: 02.06.2023
	Page no: 1 of 2

Customer details:	Reference used:
Name: Minu Clinical Laboratory	Master Universal Calibrator UC12 with RTD Sensor
Address Shop No 7/8, Sunflower, Royal complex, Eksar Road, Borivali West Mumbai - 400091	Master Ser. No. 20030512/202 ID NO: TM031/TM015
	Traceable Certificate M-211019-27-2 Godrej CC-2634
	Valid upto 21.10.2022

Details of DUC:	Condition of item: Good
Equipment: Digital Thermometer	Range °C: -50 to 300
Make: Mextech	Accuracy ±°C : 1
Model: ST - 9283B	Resolution °C: 0.1
Serial No Not Specified	Length of sensor mm: 150
ID No: Not Specified	Dia of sensor mm: 5
Location: Pahology Lab	Depth of sensor mm: 50
Ambient Temp(°C): 25.3	Humidity (%Rh): 57

Remark:

1. Calibration procedure used: LAB SOP/04
2. The Reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $K=2$, providing a level of confidence of approx. 95%.
3. The observations reported represent values at the time of measurement. They do not convey any long term stability information.
4. The standards maintained are traceable to National / International Standards through accredited laboratories.
5. Sticker affixed on the unit

Note:

1. This certificate refer only to particular instrument submitted for calibration.
2. This certificate shall not be reproduced.
3. Technical Resources shall not be responsible for any change in calibration date and performance specification on account of malfunctioning of the instrument due to damage caused to it after handing over it to the customer.

Authorised signatory
Sr Technical Manager
N Srinivas Rao

