

## CALIBRATION REPORT

Customer Name : Lotus Diagnostics

Address : Shop No. 5, Chandresh CHS,  
Next to Lotus Multi Speciality Hospital,  
Opp. ST Lawrence School, Devki Nagar,  
Borivali -West, Mumbai -400103

Instrument Model : ST-200 Plus Electrolyte Analyzer

Serial no : ST -3789

Calibration Date : 12.06.2017

There is no due date for calibration. Since calibration is an automatic process and is done every 8 hourly.



Makarand Thakur  
(Service Engineer)

## Operating the Analyzer

### 3. Operating the Analyzer

#### A. calibration

The instrument has to be calibrated when it is switched ON or it automatically calibrates on set calibrated frequency after four hours or eight hours as per set frequency.

Following is the calibration procedure:

#### A. Calibration

