

LOTUS DIAGNOSTICS (562)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Analyte	Instrument	Result Value	Standard Unit
# Blood (urine)	Manual	2+	-
# Nitrite (urine)	Manual	Absent	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

LOTUS DIAGNOSTICS (562)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01


Report Date : 30/12/2021

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5,6,6.5
@ Specific Gravity	-	1.015	1.015,1.020,1.025
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,Tr
@ Ketone (urine)	-	Negative	Negative
# Blood (urine)	-	2+	
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
# Nitrite (urine)	-	Absent	

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	2
Evaluated Parameters	7
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report



To Whomesoever it may concern

This is to certify that Urine routine test in clinical pathology is done by Urine Strip test (Manual method).

So IQ, OQ, PQ and calibration is not applicable for the same.

Janhvi

Dr Janhvi Ahuja

Laboratory Director

DR. JANHVI AHUJA
CONSULTANT PATHOLOGIST
2003/03/897



TO WHOM SO EVER IT MAY CONCERN

This is to bring to your notice that as per the scheme schedule for Clinical pathology (Urine) MHL EQAS sample is Twice a year, So we have not recieved Clinical Pathology (Urine) sample in May 2022.

Therefore we have attached EQAS result of last cycle (December 2021)

The next cycle of Clinical Pathology(Urine) sample may recieve in the December 2022 cycle.

DR. JANHVI AHUJA
CONSULTANT PATHOLOGIST
2003/03/897

Janhvi
28/06/22
Dr Janhvi Ahuja

Laboratory Director