



Date : 26 May'22

**U-Lab at Medicare Hospital, Indore
(A unit of Lahoti Medicare Pvt. Ltd.)**

Dear Sir,
As desired by you please find the Calibration report.

1. Year of manufacturing - 2000 EasyLyte Sr. No. 45423

2. Date of Calibration - 13 May'22

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	OK/not OK
Logic 5 V	5.06V	OK
Line voltage	230V	OK
Logic 12V	12V	OK
Calibration Value (Na)	59.45	OK
Calibration Value (K)	59.71	OK
Calibration Value (Cl)	47.20	OK
Electrode Values(Na)	131	OK
Electrode Value(K)	85	OK
Electrode Value(Cl)	57	OK
Pump calibration		OK
Fluid Test		OK

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For **TRANSASIA BIO-MEDICALS LTD.**



June 2022

Diagnosis	Sodium (Na)		Potassium (K)		Sign.
	N	P	N	P	
Range	129 - 145	136 - 153	3.45 - 4.13	6.08 - 7.28	
STD	137	144	3.79	6.68	
Date:-					
1/6/22	136	145	3.5	6.7	@Jadhav
2/6/22	137	143	3.6	6.6	@Jadhav
3/6/22	135	144	3.7	6.5	@Jadhav
4/6/22	137	145	3.8	6.4	meat
6/6/22	136	143	3.7	6.7	@Jadhav
7/6/22	139	142	3.5	6.3	meat
8/6/22	140	147	3.1	6.1	meat
9/6/22	137	145	3.6	6.5	@Jadhav
10/6/22	136	142	3.8	6.7	@Jadhav
11/6/22	138	141	3.9	6.5	meat
13/6/22	137	142	3.7	6.7	@Jadhav
14/6/22	136	145	3.9	6.1	meat
15/6/22	135	146	3.7	6.7	@Jadhav
16/6/22	137	143	3.6	6.8	@Jadhav
17/6/22	135	140	3.2	5.9	meat
18/6/22	136	145	3.1	6.0	meat
20/6/22	137	145	3.8	6.7	@Jadhav
21/6/22	136	144	3.7	6.7	@Jadhav
23/6/22	136	148	3.8	6.6	@Jadhav
23/6/22	139	149	3.9	5.8	meat
24/6/22	142	149	4.2	6.7	meat
25/6/22	136	144	3.8	6.7	@Jadhav
27/6/22	137	146	3.7	6.8	@Jadhav
28/6/22	139	149	3.8	6.6	meat
29/6/22	138	148 148	3.9	6.4	meat