

CALIBRATION CERTIFICATE

For : U-Lab At Medicare Hospital (A Unit Of Lahoti Medicare Pvt. Ltd.), Indore

Calibration Certificate No. : LMPL/INDORE/05/22/REFRIGERATOR-1		Instrument Code No. : REFRIGERATOR-1	
Instrument Name	: REFRIGERATOR 2-8	Calibration Date	: 15.05.2022
Section	: PATHOLOGY LABORATORY	Calibration Due Date	: 14.05.2023
Location	: PATHOLOGY LABORATORY		
Make & Model	: GODREJ	Range	: 5 ± 3 °C
Serial No.	: ----	Specified Accuracy	: ----
Least Count	: 0.1 °C	Acceptance Criteria	: ± 1 °C
Environment Condition :		Calibration SOP NO. : QATECH/CAL - 014	
Temperature : 24.5 °C		Relative Humidity : 57.2 % RH	

OBSERVATION TABLE :

Sr. No.	Input / Cal. Point °C	Reading Before Adjustment			Reading After Adjustment			Remark
		Master Reading °C	UUC Reading °C	Deviation °C	Master Reading °C	UUC Reading °C	Deviation °C	
1	3.0	3.2	3.5	0.3	----	----	----	OK


CALIBRATION STANDARDS USED :

Name & Make	Range	Accuracy	Certificate No.	Calibration Valid Up To
DIGITAL THERMOMETER WITH SENSOR, MAKE : UDIAN, SR. NO.: 201550424, ID. NO.: QA/MUM/CAL/DTS-18 (INDICATOR) QA/MUM/CAL/DTS-18A,18B (SENSOR)	-100 TO 300 °C	AS PER MANUAL	QATM - 22021407	22.02.2023

REMARK :

- The refrigerator 2-8 REFRIGERATOR-1 is calibrated within acceptance criteria limit.
- This certificate refers only to the particular item calibrated at Site.
- The calibration results reported are valid at the time of calibration & under the condition of environment.
- The calibration standard used are traceable to National / International Standards.
- 'UUC' Unit Under Calibration.
- Calibration is done for points, selected in working range as per user requirement.

Note : This certificate should not be reproduced, except in full, without the prior permission of QA Tech.

 MS SIGN.	MS
Calibration Certificate Issued By:	Calibration Certificate Checked By:

* End of Calibration Certificate *