

CALIBRATION CERTIFICATE

For : U-Lab At Medicare Hospital (A Unit Of Lahoti Medicare Pvt. Ltd.), Indore

Calibration Certificate No. : LMPL/INDORE/05/22/CENTRIFUGE-1	Instrument Code No. : CENTRIFUGE-1
Instrument Name : CENTRIFUGE	Calibration Date : 15.05.2022
Section : PATHOLOGY LABORATORY	Calibration Due Date : 14.05.2023
Location : PATHOLOGY LABORATORY	
Make & Model : REMI & R-8C	Range : MAX 5250 RPM
Serial No. : ----	Specified Accuracy : ----
Least Count : 1 RPM	Acceptance Criteria : $\pm 2\%$ F.S.
Environment Condition : Temperature : 24.8 °C	Calibration SOP NO. : QATECH/CAL - 020 Relative Humidity : 56.4 % RH

OBSERVATION TABLE :

Sr. No.	Input / Cal. Point RPM	Reading Before Adjustment			Reading After Adjustment			Remark
		Master Reading RPM	UUC Reading RPM	Deviation RPM	Master Reading RPM	UUC Reading RPM	Deviation RPM	
1	1000	996.5	1000	3.5	----	----	----	OK
2	3000	2995.3	3000	4.7	----	----	----	OK

CALIBRATION STANDARDS USED :

Name & Make	Range	Accuracy	Certificate No.	Calibration Valid Up To
DIGITAL TACHOMETER, MAKE : MONARCH, MODEL : PLT-200, SR. NO.: 1846187, ID. NO.: QA/MUM/CAL/PLT-06	5 To 200000 RPM	AS PER MANUAL	QATM - 22020627	04.02.2023

REMARK :

- The centrifuge CENTRIFUGE-1 is calibrated within acceptance criteria limit.
- This certificate refers only to the particular item calibrated at Site.
- The calibration results reported are valid at the time of calibration & under the condition of environment.
- The calibration standard used are traceable to National / International Standards.
- 'UUC' Unit Under Calibration.
- Calibration is done for points selected in working ranges as per user requirement.

Note : This certificate should not be reproduced, except in full, without the prior permission of QA Tech.

 Calibration Certificate Issued By: SIGN	MS Calibration Certificate Checked By:
---	---

* End of Calibration Certificate *