



Maruthi Enterprises

103, 4th Main, Gangappa Block, Ganganagar, Bangalore - 560 032.
E-mail : maruthi.enterprises@yahoo.com

CALIBRATION CERTIFICATE

CUSTOMER:- REHA DIAGNOSTIC LABORATORY,
#15/A, OLD POLICE STATION ROAD,
KONANAKUNTE,
BANGALORE-62.

CERTIFICATE NO :- MEC -8976
CALIBRATION DATE :- 01/01/2022

DESCRIPTION OF ITEM :- LAB BIOCHEMISTRY ANALYZER

MODEL :- RX-50V

S.NO :- 1844153

RANGE :- 8 FILTERS

DETAILS OF EQUIPMENT USED FOR CALIBRATION :-

Analysis method.

Absorbance, Endpoint, Fixed time, Kinetic, 2-point end, Bichromatic & Multi standard.

CALIBRATION DATA :-


S.NO	CALIBRATION	DISPLAY	OBSERVED	REMARKS
	OPTICAL SYSTEM	READING IN OPTICAL	READING BY	
1	8 FILTERS	340 nm	32 ul quartz flow cell	ok
2	8 FILTERS	405 nm	4 temperature option	ok
3	8 FILTERS	450 nm	25' c, 30' c, 37'c	ok
4	8 FILTERS	505 nm	range from 0.0000-	ok
5	8 FILTERS	546 nm	3.0000 abs	ok
6	8 FILTERS	578 nm	resolution -0.0001 abs	ok
7	8 FILTERS	630 nm	stability -0.005 abs/h	ok
8	8 FILTERS	700 nm	light - 6v/10w holgen b	ok

CALIBRATION DUE ON:- 01/01/2023

NOTE:-

1. THIS CALIBRATION CERTIFICATE RELATES ONLY TO THE ABOVE DUC.
2. CALIBRATION CERTIFICATE SHALL NOT BE REPRODUCED EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL FROM BCL.
3. RESULT REPORTED ARE VALID AT THE TIME OF AND UNDER THE STATED CONDITION OF MEASUREMENT.
4. ANY ERROR IN THIS REPORT SHOULD BE BROUGHT TO OUR KNOWLEDGE WITHIN 45 DAYS FROM THE DATE OF THIS CERTIFICATE.

CALIBRATED BY :-
MURTHY B.



CHECKED BY:-
DRISHYA K.K.