



Maruthi Enterprises

103, 4th Main, Gangappa Block, Ganganagar, Bangalore - 560 032.

E-mail : maruthi.enterprises@yahoo.com

CALIBRATION CERTIFICATE

CUSTOMER:- REHA DIAGNOSTIC LABORATORY,
#15/A, OLD POLICE STATION ROAD,
KONANAKUNTE,
BANGALORE-62.

CERTIFICATE NO :- MEC -8979
CALIBRATION DATE :- 01/01/2022

DESCRIPTION OF ITEM :- LAB AUTO HEMATOLOGY ANALYZER

MODEL :- H 33S

S.NO :- TH -83004526

RANGE :- 0.00 TO 150.00'C

OUT PUT RANGE :- 4.00 TO 20.00 Ma

DETAILS OF EQUIPMENT USED FOR CALIBRATION :-

Reject Error :- 0.50% of Range.

Temperature :- 22.0'C

Relative Humidity :- 50%

Device method.

TAG/ID: TMP -TX, TEMPERATURE TRANSMITTER, 0-150'C 4-20mA.

CALIBRATION DATA :-

S.NO	CALIBRATION TEST POINT	DISPLAY INPUT TEMPERATURE	OBSERVED NOMINAL OUTPUT	REMARKS ALLOWED
1	0%	0.00'C	4.00MA	0.1MA
2	20%	30.00'C	7.20MA	0.1MA
3	40%	60.00'C	10.40MA	0.1MA
4	60%	90.00'C	13.63MA	0.1MA
5	80%	120.00'C	16.80MA	0.1MA
6	100%	150.00'C	20.02MA	0.1MA

CALIBRTION BUE ON:- 01/01/2023

NOTE:-

1. THIS CALIBRTION CERTIFICATE RELATES ONLY TO THE ABOVE DUC.
2. CALIBRTION CERTIFICATE SHALL NOT BE REPRODUCED EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVED FROM BCL.
3. RESULT REPORTED ARE VALID AT THE TIME OF AND UNDER THE STATED CONDTIONOF MEASUREMENT.
4. ANY ERROR IN THIS REPORT SHOULD BD BROUGHT TO OUR KNOWLEDGE WITHIN 45 DAYS FROM THE DATE OF THIS CERTIFICATE.

CALIBRATED BY :-
MURTHY B.



CHECKED BY:-
DRISHYA K.K.