

# Calibration Certificate

Cell : 99804 08580



## Maruthi Enterprises

# 103, 4th Main, Gangappa Block, Ganganagar, Bangalore - 560 032.  
E-mail : maruthi.enterprises@yahoo.com

### CALIBRATION CERTIFICATE

CUSTOMER:- REHA DIAGNOSTIC LABORATORY,  
#15/A, OLD POLICE STATION ROAD,  
KONANAKUNTE, BANGALORE-62.

CERTIFICATE NO :- MEC - 3456

CALIBRATION DATE :- 01/01/2022

DESCRIPTION OF ITEM :- REMI CENTRIFUGE

MODEL :- C-852/6

S.NO :- REMI -01

RANGE :- 350 TO 3500 RPM

DETAILS OF EQUIPMENT USED FOR CALIBRATION :-

S.NO	DESCRIPTION	CERTIFIED BY	CERTIFICATE NO	VALIDITY
1	DIGITAL NON CONTACT TACHOMETER	BCL	SRF NO:- 4375 - 01	2/21/2022

CALIBRATION DATA :-


S.NO	CALIBRATION	DISPLAY	OBSERVED	REMARKS
	POINTS(IN RPM)	READING	ERROR	
1	1500 RPM	1450 RPM	50 RPM	OK
2	2500 RPM	2470 RPM	30 RPM	OK
3	3500 RPM	3560 RPM	60 RPM	OK

NEXT CALIBRATION :- 01/01/2023

NOTE:-

1. THIS CALIBRTION CERTIFICATE RELATES ONLY TO THE ABOVE DUC.
2. CALIBRTION CERTIFICATE SHALL NOT BE REPRODUCED EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVED FROM BCL.
3. RESULT REPORTED ARE VALID AT THE TIME OF AND UNDER THE STATED CONDTIONOF MEASUREMENT.
4. ANY ERROR IN THIS REPORT SHOULD BD BROUGHT TO OUR KNOWLEDGE WITHIN 45 DAYS FROM THE DATE OF THIS CERTIFICATE.

CALIBRATED BY :-  
MURTHY B.



CHECKED BY:-  
DRISHYA K.K.