

# Calibration Certificate

Cell : 99804 08580



## Maruthi Enterprises

# 103, 4th Main, Gangappa Block, Ganganagar, Bangalore - 560 032.  
E-mail : maruthi.enterprises@yahoo.com

### CALIBRATION CERTIFICATE

CUSTOMER:- REHA DIAGNOSTIC LABORATORY,  
#15/A, OLD POLICE STATION ROAD,  
KONANAKUNTE,  
BANGALORE-62.

CERTIFICATE NO :- MEC -21246  
CALIBRATION DATE :- 01/01/2022

DESCRIPTION OF ITEM :- REMI CENTRIFUGE

MODEL :- R-8C/16 TUBES

S.NO :- ZEEN -18118

RANGE :- 350 TO 5500 RPM

DETAILS OF EQUIPMENT USED FOR CALIBRATION :-

| S.NO | DESCRIPTION                       | CERTIFIED BY | CERTIFICATE NO    | VALIDITY   |
|------|-----------------------------------|--------------|-------------------|------------|
| 1    | DIGITAL NON CONTACT<br>TACHOMETER | BCL          | SRF NO:- 4375 -01 | 21/02/2022 |

CALIBRATION DATA :-

| S.NO | CALIBRATION<br>POINTS(IN RPM) | DISPLAY<br>READING | OBSERVED<br>ERROR | REMARKS |
|------|-------------------------------|--------------------|-------------------|---------|
| 1    | 1500 RPM                      | 1450 RPM           | 50 RPM            | OK      |
| 2    | 2500 RPM                      | 2470 RPM           | 30 RPM            | OK      |
| 3    | 3500 RPM                      | 3560 RPM           | 60 RPM            | OK      |
| 4    | 4500 RPM                      | 4530 RPM           | 30 RPM            | OK      |

T CALIBRTION BUE ON:- 01/01/2023

NOTE:-

1. THIS CALIBRTION CERTIFICATE RELATES ONLY TO THE ABOVE DUC
2. CALIBRTION CERTIFICATE SHALL NOT BE REPRODUCED EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVED FROM BCL.
3. RESULT REPORTED ARE VALID AT THE TIME OF AND UNDER THE STATED CONDITONOF MEASUREMENT.
4. ANY ERROR IN THIS REPORT SHOULD BD BROUGHT TO OUR KNOWLEDGE WITHIN 45 DAYS FROM THE DATE OF THIS CERTIFICATE.

CALIBRATED BY :-  
MURTHY B.



CHECKED BY:-  
DRISHYA R.K.