

SERVICE REPORT

047691



Customer ID	CUST 1909	Customer Designation	OWNER		
Call Log No.	BRKDOWN123364	Date	30 APRIL 2022		
Name of the Lab	RENUKA DIAGNOSTICS PATHOLOGY				
Name of the Customer	DR. AMIT MISHRA				
Address	NEAR MAHARANA PRATAP CHOWK, PINGU ROAD-2, GAURAV PATH, BILASPUR, C.G.				
Land Mark	MAHILA GRAMIN BANK	Pin Code	495001		
Mobile No.	9098161081	Land Line No.	NA		
E Mail Address	amitmishra552@gmail.com				
Instrument Name	CG-MICRO	Serial No.	CM0921150		
Date of Installation	14 JAN 2022	IW/OW /AMC /CMC/AQ	IW		
Call Type (RR/PRR/GS/Out right/AQ) - (Install/Breakdown/Backup/Training/PM/Reagent)	BREAKDOWN				
Reagent Catalogue No.	NA	PQCF Date	NA		
Reagent Lot No.	NA	Expiry Date	NA		
Calibration / Std Lot No.	NA	Expiry Date	NA		
Call Attended On	30 APRIL 2022	Service Time	2 HRS.		
Special Remark from CC	NA	Travel Time	2 HRS.		
Problem Reported	POWER SUPPLY ISSUE	Actual Problem	ADAPTOR IS FAULTY		
Action on Problem	CHANGED THE ADAPTOR WITH NEW ONE AND SUBMITTING THE OLD ONE TO THE O.H.O. (MUMBAI) ALSO ADJUSTED THE GAIN OF THE FILTERS				
Spares Replaced / Required to Replace					
Part No.	Part Name	Invoice No.	Invoice Date	Price	Qty
	ADAPTOR				1
Reagent Consumption by Engineer on the Machine					
Name of Reagent	Test Conducted	Standard Consumption	Actual Consumption	Remarks	
NA	NA	NA	NA	NA	
NA	NA	NA	NA	NA	
NA	NA	NA	NA	NA	
The above Merilyzer is Repaired <input checked="" type="checkbox"/> / Serviced <input type="checkbox"/> / Demonstrated <input type="checkbox"/> / Trained User <input type="checkbox"/> and is working satisfactorily					
Engineer / Application Remark	INSTRUMENT IS WORKING FINE	Customer's Remark	Thanking All best		
Signature of Engr / App Name of Engr / App Date	Mishra MAYANK 30 APRIL 2022	Signature and Seal of Customer			