



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MARCH 2022



PC-1024

Lab Name MEDQUEST HELTHCARE Lab No 15856
 Constituent Group HbA1c Date of Result Entered : 20/03/2022
 PT item Lyophilized Whole blood Date of Report Published : 06/04/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	89	6.84	9.66	0.66	5.9 %	-1.42	0.14

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 MEDQUEST HELTHCARE
 DOOR NO,25-2-134 ,134 (1) &134 (2), MAHIM COMPLEX, KANKANADY OLD ROAD , KANKANADY MANGALORE,
 KARNATAKA
 KARNATAKA575002

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****