

Mindray BS-240 Auto Chemistry Analyzer – PQ Documents

Customer Address: **Mediquest Health care Diagnostic**
Mangalore -575002

Contact person: **Ms.Deeksha**
Telephone: 7760804801

Serial no: **YX-91002366**

QC Running: 06/06/22

Performance Qualification

CALIBRATION:

The Analyser has been calibrated with Mindray Multi Sera calibrator for Calcium, Glucose-G, T.Bilirubin, D.Bilirubin, ALT, AST, GGT, ALP, GGT, TP, TGL, HDL-C, TC, UA, Crea-S.

Calibrator Manufacturer : Mindray

Lipids Calibrator : Mindray

Calibrator name : Multi Sera calibrator

Calibrator Name : LIPID CAL

Lot Number : 150121003

Lot Number : 150421013

Expiry date : 18-10-2022

Expiry Date : 30-11-2022

QUALITY CONTROL CHECK - NORMAL CONTROL

Control manufacturer : RANDOX

QC Name : Randox L2 & L3

Lot Number : 1502UN & 1173UE

Expiry Date : 28-06-2024 & 28-06-2024

Carried out Randox Control found values are within specified range.

REMARKS

The system meets all criteria outlined in this performance qualification documents. All the required parameter have been checked and instrument is ready for regular use:

Performer : **Viswanatha. T**
Designation : Application specialist
Date : 21/06/2022

Reviewer : **Muthu vigneshwar. K.P**
Designation : Application Manager

