MK BEST CALIBRATION SERVICES



एमके बेस्ट केलिब्रेशन सर्विसेस्

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal, (Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



Date of Issue Certificate No MKBN/22/07/0608-001 00 07 0000 Basam Dua Data

Recom.D	ue Date : 07.07	9001:2015	
U	NIT UNDER CALIBRATION		
Description	LABORATORY CENTR	RIFUGE	
VT Range	5250 rpm 100 rpm		
Resolution			
Make			
Model R - 8C			
Identification No	SSL/TVR/EQP/CB/001		
Serial No	ZHEN-15036		
Manufacturer Name	REMI		
Traceability	Certificate No	Validity	
tional Standards	CRMTL/01/421103387-A5	20.10.2022	
ENVIRONMENTAL CONDITIONS REFERENCE STANDARD & ACCEPTANCE LIMIT			
	Description VT Range Resolution Make Model Identification No Serial No Manufacturer Name Traceability tional Standards	UNIT UNDER CALIBRATION Description LABORATORY CENTER Traceability UNIT UNDER CALIBRATION LABORATORY CENTER 5250 rpm 100 rpm 100 rpm Make Nodel R - 8C SSL/TVR/EQP/CB/O Serial No ZHEN-15036 REMI Traceability Certificate No CRMTL/01/421103387-A5	

CALIBRATION RESULTS

Uncertainty ± °C

S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	507	500	-7
2	1015	1000	-15
3	2022	2000	-22
4	3027	3000	-27
5	4532	4500	-32
6	5241	5200	-41

Remarks:

Temperature

Humidity

20.4°C

52 % RH

1. The Expanded Uncertainty associated with the results are calculated at a confidence level of approximately 95% with a Coverage factor k=2.

2. The above UUC was Calibrated at Site.

Calibrated by

S.Chandra Bose

(Calibration Engineer)

Authorised By

1.57

L.Magesh (MD/QM)

THE STANDARD INSTRUMENTS USED ARE TRACEABLE TO NABL