

CALIBRATION CERTIFICATE

For : U-Lab At Medicare Hospital (A Unit Of Lahoti Medicare Pvt. Ltd.), Indore

Calibration Certificate No. : LMPL/INDORE/07/22/REFRIGERATOR-1A	Instrument Code No. : REFRIGERATOR-1A
Instrument Name : REFRIGERATOR (Deep freezer)	Calibration Date : 08.07.2022
Section : PATHOLOGY LABORATORY	Calibration Due Date : 07.07.2023
Location : PATHOLOGY LABORATORY	
Make & Model : GODREJ	Range : -6 ± 3 °C
Serial No. : -----	Specified Accuracy : -----
Least Count : 0.1 °C	Acceptance Criteria : ± 1 °C
Environment Condition : Temperature : 24.5 °C	Calibration SOP NO. : QATECH/CAL - 014 Relative Humidity : 57.2 % RH

OBSERVATION TABLE :

Sr. No.	Input / Cal. Point °C	Reading Before Adjustment			Reading After Adjustment			Remark
		Master Reading °C	UUC Reading °C	Deviation °C	Master Reading °C	UUC Reading °C	Deviation °C	
1	-6	-6.3	-6.1	0.2	-----	-----	-----	OK

CALIBRATION STANDARDS USED :

Name & Make	Range	Accuracy	Certificate No.	Calibration Valid Up To
DIGITAL THERMOMETER WITH SENSOR, MAKE : UDIAN, SR. NO.: 201550424, ID. NO.: QA/MUM/CAL/DTS-18 (INDICATOR) QA/MUM/CAL/DTS-18A,18B (SENSOR)	-100 To 300 °C	AS PER MANUAL	QATM - 22021407	22.02.2023

REMARK :

- The refrigerator (deep freezer) REFRIGERATOR-1A is calibrated within acceptance criteria limit.
- This certificate refers only to the particular item calibrated at Site.
- The calibration results reported are valid at the time of calibration & under the condition of environment.
- The calibration standard used are traceable to National / International Standards.
- 'UUC' Unit Under Calibration.
- Calibration is done for points, selected in working range as per user requirement.

Note : This certificate should not be reproduced, except in full, without the prior permission of QA Tech.

Calibration Certificate Issued By:	MS
Calibration Certificate Checked By:	

* End of Calibration Certificate *