

LABORATORY VILLOOPOONAWALLA MEMORIAL HOSPITAL (637)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Negative	Neg

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5	5,5,5,6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	4+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,Tr
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	2+	+1,+2,Tr
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Absent	Abs

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	10
Not Evaluated Parameters	0
Evaluated Parameters	10
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416,  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurda (W),  
 Mumbai - 400070

End of Report

*Reviewed results.  
 Satisfactory.  
 Hllah  
 8/7/2022*