

Instrument : Dialab Analyser

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	66.50	--	0.24	0.37
@ Urea	mg/dl	21.70	--	1.01	0.25
@ Creatinine	mg/dl	0.68	--	-0.20	-0.60
@ Bilirubin Direct	mg/dl	0.16	--	1.71	-0.50
@ Total Protein	g/dL	6.77	--	1.34	0.73
# Albumin	g/dL	3.12	--	-2.16	-2.09
@ ALP	U/L	63.50	--	1.23	1.67
@ CK	U/L	50.10	--	0.98	0.98
@ GGT	U/L	23.10	--	-0.19	-0.10
@ LDH	U/L	144.00	--	-1.53	-2.33
@ Uric Acid	mg/dl	3.49	--	-0.91	-0.89
# Cholesterol	mg/dl	123.00	--	-2.21	-1.12
# Triglycerides	mg/dl	151.00	--	2.87	2.11
X HDL	mg/dl	53.70	--	3.70	2.84
X ALT	U/L	11.60	--	3.80	5.53
@ AST	U/L	18.20	--	0.00	0.21

Instrument : Iris IONS

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
X Chloride	mmol/L	189.30	--	6.84	6.84
@ Sodium	mmol/L	128.40	--	-0.02	-0.02
@ Potassium	mmol/L	3.99	--	0.82	0.57

Instrument : MicroLab RX-50

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Calcium	mg/dl	9.00	--	1.15	1.12
@ Phosphorous Inorganic	mg/dl	5.72	--	1.84	1.54
@ Magnesium	mg/dl	1.70	--	-0.16	-0.23
@ Amylase	U/L	52.50	--	-1.80	-1.18
@ Lipase	U/L	32.19	--	-0.94	-0.67

Legend @ : z score \leq 2.0 - Acceptable
 # : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	24
Not Evaluated Parameters	0
Evaluated Parameters	24
Outlier Parameters (X)	3
EQAS Score Biochemistry	87.50 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

Reviewed EQAS results.
 Satisfactory.
Subhakar