

**Villoo Poonawalla Memorial Hospital**  
 S.No.156, Plot No 1/3A+3B/1+2, 3, Pune - Solapur Rd,  
 Near Savli Corner, Hadapsar, Pune - 411028

Sr. No.	Instrument Name	Type	ID No.	Make	Range	Unit	L.C.	Cal. Date	Due Date
1	Thermohygrometer	Digital	VPMH/THY/001	HTC-1	-50 to 70°C/10 to 99%RH		0.1 °C/ 1 %RH	24.03.2022	23.03.2023
2	Thermohygrometer	Digital	VPMH/THY/002	HTC-1	-50 to 70°C/10 to 99%RH		0.1 °C/ 1 %RH	24.03.2022	23.03.2023
3	Thermohygrometer	Digital	VPMH/THY/003	HTC-1	-50 to 70°C/10 to 99%RH		0.1 °C/ 1 %RH	24.03.2022	23.03.2023
4	Thermohygrometer	Digital	VPMH/THY/004	HTC-1	-50 to 70°C/10 to 99%RH		0.1 °C/ 1 %RH	24.03.2022	23.03.2023
5	Thermohygrometer	Digital	VPMH/THY/005	HTC-1	-50 to 70°C/10 to 99%RH		0.1 °C/ 1 %RH	24.03.2022	23.03.2023
6	Thermometer With Sensor	Digital	VPMH/VTI/TM/559	Mult/ST-9283 B	-50 to 300	°C	0.1	24.03.2022	23.03.2023
7	Thermometer With Sensor	Digital	VPMH/VTI/TM/558	Mult/ST-9283 B	-50 to 300	°C	0.1	24.03.2022	23.03.2023



# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1. CUSTOMER</b>	:-	Page No.	:- 1 of 1
<b>Viloo Poonawalla Memorial Hospital</b>		NABL Accreditation No.	:- CC-2927
S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd,		Certificate No.	:- 22.03.24.001
Near Savli Corner, Hadapsar, Pune - 411028		Date of Received	:- 24.03.2022
Temperature(°C)	:- 23.3	Date of Calibration	:- 24.03.2022
Relative Humidity (%RH)	:- 51	Next Calibration Due On	:- 23.03.2023
Location of calibration	:- In Lab	Calibration method No.	:- RTS-WI-25
Condition of Item	:- Ok	Date of Issue	:- 26.03.2022
		ULR No.	:- CC292722000002866F

### 2. Description of Item

Name	:- Thermohygrometer	Range	:- -50 to 70°C/10 to 99%RH
Id No	:- VPMH/THY/001	Least Count	:- 0.1 °C/ 1 %RH
Make/Model	:- HTC-1		
Type	:- Digital		

### 3. Detail of Equipment used for calibration

Name	:- Dig. Thermo Hygrometer with sensor	6½ DMM	PT 100 Sensor
Certificate No.	:- 30045622	VI/21-22/0472-02	TL/021/531.1.2
Certified By	:- Belz	VIPL	Tempsens
ID/Sr. No.	:- 61319921	TW00006839	19640
Range	:- -40 to 60 °C/ 0 to 100% RH	As per traceability	-200 to 500 °C
Uncertainty	:- ± 0.2 °C/ ± 0.8% RH	As per traceability	± 0.15 °C
Calibration Validity	:- 28.10.2022	28.09.2022	10.06.2022

Discipline :- Thermal Calibration

Group :- Specific heat & Humidity

### 4. Calibration Results

#### A) Temperature

Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
10	10.03	10.1	0.07	0.30
30	30.08	30.2	0.12	0.30
50	50.14	50.4	0.26	0.30

#### B) Humidity


Calibration Points %RH	Standard Reading %RH	UUC Reading %RH	Error in %RH	Expanded Uncertainty in ± %RH
20	20.1	21	0.9	1.62
70	70.6	72	1.4	1.62
90	91.3	93	1.7	1.62

#### Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- 6) The Standard used are traceable to National / International Standard.
- 7) Conformity with requirements :-

Calibrated By

*Divya M*  
Calibration Engineer  
Divya Dagdu



Approved By

*SM*  
Technical Manager  
Sachin A. Mhasawade

RF-21, R0







# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834 35, Sant Tukaram Nagar,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170 .  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Viloo Poonawalla Memorial Hospital</b>	NABL Accreditation No.	:- CC-2927
S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd,	Certificate No.	:- 22.03.24.002
Near Savli Corner, Hadapsar, Pune - 411028	Date of Received	:- 24.03.2022
Temperature(°C) :- 22.4	Date of Calibration	:- 24.03.2022
Relative Humidity (%RH) :- 51	Next Calibration Due On	:- 23.03.2023
Location of calibration :- In Lab	Calibration method No.	:- RTS-WI-25
Condition of Item :- Ok	Date of Issue	:- 26.03.2022
	ULR No.	:- CC292722000002867F

### 2. Description of Item

Name :- Thermohygrometer	Range :-	-50 to 70°C/10 to 99%RH
Id No :- VPMH/THY/002	Least Count :-	0.1 °C/ 1 %RH
Make/Model :- HTC-1		
Type :- Digital		

### 3.Detail of Equipment used for calibration

Name :- Dig. Thermo Hygrometer with sensor	6½ DMM	PT 100 Sensor
Certificate No. :- 30045622	VI/21-22/0472-02	TL/021/531.1.2
Certified By :- Belz	VIPL	Tempsens
ID/Sr. No. :- 61319921	TW00006839	19640
Range :- -40 to 60 °C/ 0 to 100% RH	As per traceability	-200 to 500 °C
Uncertainty :- ± 0.2 °C/ ± 0.8% RH	As per traceability	± 0.15 °C
Calibration Validity :- 28.10.2022	28.09.2022	10.06.2022

Discipline :- Thermal Calibration

Group :- Specific heat & Humidity

### 4. Calibration Results :-

#### A) Temperature

Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
10	10.03	10.2	0.17	0.30
30	30.08	30.3	0.22	0.30
50	50.14	50.5	0.36	0.30

#### B) Humidity

Calibration Points %RH	Standard Reading %RH	UUC Reading %RH	Error in %RH	Expanded Uncertainty in ± %RH
20	20.1	20	-0.1	1.62
70	70.6	69	-1.6	1.62
90	91.3	89	-2.3	1.62

#### Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- The Standard used are traceable to National / International Standard.
- Conformity with requirements :- ----

Calibrated By

*Divya*  
Calibration Engineer  
Divya Dagdu



Approved By

*SM*  
Technical Manager  
Sachin A. Mhasawade







# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Viloo Poonawalla Memorial Hospital</b>	NABL Accreditation No.	:- CC-2927
S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd,	Certificate No.	:- 22.03.24.003
Near Savli Corner, Hadapsar, Pune - 411028	Date of Received	:- 24.03.2022
Temperature(°C) :- 22.6	Date of Calibration	:- 24.03.2022
Relative Humidity (%RH) :- 51	Next Calibration Due On	:- 23.03.2023
Location of calibration :- In Lab	Calibration method No.	:- RTS-WI-25
Condition of Item :- Ok	Date of Issue	:- 26.03.2022
	ULR No.	:- CC292722000002868F

### 2. Description of Item

Name :- Thermohygrometer	Range	:- -50 to 70°C/10 to 99%RH
Id No :- VPMH/THY/003	Least Count	:- 0.1 °C/ 1 %RH
Make/Model :- HTC-1		
Type :- Digital		

### 3.Detail of Equipment used for calibration

Name :- Dig. Thermo Hygrometer with sensor	C $\frac{1}{2}$ DMM	PT 100 Sensor
Certificate No. :- 30045622	VI/21-22/0472-02	TL/021/531.1.2
Certified By :- Belz	VIPL	Tempsens
ID/Sr. No. :- 61319921	TW000006839	19610
Range :- -40 to 60 °C/ 0 to 100% RH	As per traceability	-200 to 500 °C
Uncertainty :- $\pm 0.2$ °C/ $\pm 0.8\%$ RH	As per traceability	$\pm 0.15$ °C
Calibration Validity :- 28.10.2022	28.09.2022	10.06.2022

Discipline :- Thermal Calibration Group :- Specific heat & Humidity

### 4. Calibration Results :-

#### A) Temperature

Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in $\pm$ °C
10	10.12	10.2	0.08	0.30
30	30.31	30.5	0.19	0.30
50	50.36	50.7	0.34	0.30

#### B) Humidity


Calibration Points %RH	Standard Reading %RH	UUC Reading %RH	Error in %RH	Expanded Uncertainty in $\pm$ %RH
20	20.2	19	-1.2	1.62
70	70.7	69	-1.7	1.62
90	91.4	89	-2.4	1.62

#### Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- The Standard used are traceable to National / International Standard.
- Conformity with requirements :- ----

Calibrated By

*Divya*  
Calibration Engineer  
Divya Dagdu



Approved By

*SM*  
Technical Manager  
Sachin A. Mhasawade







# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :-	Page No.	:- 1 of 1
<b>Villoo Poonawalla Memorial Hospital</b>	NABL Accreditation No.	:- CC-2927
S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd,	Certificate No.	:- 22.03.24.004
Near Savli Corner, Hadapsar, Pune - 411028	Date of Received	:- 24.03.2022
Temperature(°C) :- 23.2	Date of Calibration	:- 24.03.2022
Relative Humidity (%RH) :- 55	Next Calibration Due On	:- 23.03.2023
Location of calibration :- In Lab	Calibration method No.	:- RTS-WI-25
Condition of Item :- Ok	Date of Issue	:- 26.03.2022
	ULR No.	:- CC292722000002869F

### 2. Description of Item

Name :- Thermohygrometer	Range :- -50 to 70°C/10 to 99%RH
Id No :- VPMI/TIIV/004	Least Count 0.1 °C/ 1 %RH
Make/Model :- HTC-1	
Type :- Digital	

### 3.Detail of Equipment used for calibration

Name :- Dig.Thermo Hygrometer with sensor	6½ DMM	PT 100 Sensor
Certificate No :- 30016622	VI/21 22/0472-02	TL/021/501.1.2
Certified By :- Belz	VIPL	Tempens
ID/Sr. No. :- 61319921	TW00006839	19640
Range :- -40 to 60 °C/ 0 to 100% RH	As per traceability	-200 to 500 °C
Uncertainty :- ± 0.2 °C/ ± 0.8% RH	As per traceability	± 0.15 °C
Calibration Validity :- 28.10.2022	28.09.2022	10.06.2022

**Discipline** :- Thermal Calibration **Group** Specific heat & Humidity

### 4.Calibration Results :-

#### A) Temperature

Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
10	10.12	10.4	0.28	0.30
30	30.31	30.7	0.39	0.30
50	50.36	50.8	0.44	0.30

#### B) Humidity

Calibration Points %RH	Standard Reading %RH	UUC Reading %RH	Error in %RH	Expanded Uncertainty in ± %RH
20	20.2	21	0.8	1.62
70	70.7	72	1.3	1.62
90	91.4	94	2.6	1.62

#### Note:


- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- The Standard used are traceable to National / International Standard.
- Conformity with requirements :- ----

Calibrated By

*Divya*

Calibration Engineer

Divya Dagdu



Approved By

*S.M.*

Technical Manager

Sachin A. Mhasawade







# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE WORLD" 497720J4-J5, Sant Tukaram Nagari,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1. CUSTOMER</b>	:-	Page No.	:- 1 of 1
<b>Viloo Poonawalla Memorial Hospital</b>		NABL Accreditation No.	:- CC-2927
S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd,		Certificate No.	:- 22.03.24.005
Near Savli Corner, Hadapsar, Pune - 411028		Date of Received	:- 24.03.2022
Temperature(°C)	:- 22.4	Date of Calibration	:- 24.03.2022
Relative Humidity (%RH)	:- 51	Next Calibration Due On	:- 23.03.2023
Location of calibration	:- In Lab	Calibration method No.	:- RTS-WI-25
Condition of Item	:- Ok	Date of Issue	:- 26.03.2022
		ULR No.	:- CC292722000002870F

### 2. Description of Item

Name	:- Thermohygrometer	Range	:- -50 to 70°C/10 to 99%RH
Id No	:- VPMH/THY/005	Least Count	0.1 °C/ 1 %RH
Make/Model	:- HTC-1		
Type	:- Digital		

### 3. Detail of Equipment used for calibration

Name	:- Dig. Thermo Hygromotor with sensor	6½ DMM	PT 100 Sensor
Certificate No.	:- 30045622	VI/21-22/0472-02	TL/021/531.1.2
Certified By	:- Belz	VIPL	Tempens
ID/Sr. No.	:- 61319921	TW00006839	19640
Range	:- -40 to 60 °C/ 0 to 100% RH	As per traceability	-200 to 500 °C
Uncertainty	:- ± 0.2 °C/ ± 0.8% RH	As per traceability	± 0.15 °C
Calibration Validity	:- 28.10.2022	28.09.2022	10.06.2022

Discipline :- Thermal Calibration Group Specific heat & Humidity

### 4. Calibration Results

#### A) Temperature

Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
10	10.13	10.4	0.27	0.30
30	30.24	30.6	0.36	0.30
50	50.41	50.8	0.39	0.30

#### B) Humidity

Calibration Points %RH	Standard Reading %RH	UUC Reading %RH	Error in %RH	Expanded Uncertainty in ± %RH
20	20.3	20	0.3	1.62
70	70.8	69	-1.0	1.62
90	91.2	89	-2.2	1.62

#### Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- The Standard used are traceable to National / International Standard.
- Conformity with requirements :-

Calibrated By

*Divya Dagdu*  
Calibration Engineer  
Divya Dagdu



Approved By

*S.M.*  
Technical Manager  
Sachin A. Mhasawade







# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>Villoo Poonawalla Memorial Hospital</b> S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd, Near Savli Corner, Hadapsar, Pune - 411028 Ambient Temp. (°C) :- 23.2 Relative Humidity (%RH) :- 53 Location of calibration :- In Lab Condition of Item :- Ok	Page No. :- 1 of 1 NABL Accreditation No. :- CC-2927 Certificate No. :- 22.03.24.006 Date of Received :- 24.03.2022 Date of Calibration :- 24.03.2022 Next Calibration Due On :- 23.03.2023 Calibration method No. :- RTS-WI-10 Date of Issue :- 26.03.2022 ULR No. :- CC292722000002871F
---	---

<b>2. Description of Item</b>	
Name :- Thermomotor With Sensor	Range :- -50 to 300 °C
Id No :- VPMH/VTI/TM/559	Least Count :- 0.1 °C
Make :- Multi/ST-9283 B	
Type :- Digital	

<b>3.Detail of Equipment used for calibration</b>	
Name :- 6½ DMM	PT 100 Sensor
Certificate No. :- VI/21-22/0472-02	TL/021/531.1.2
Certified By :- VIPL	Tempens
ID/Sr. No. :- TW00006839	19640
Range :- As per traceability	-200 to 500 °C
Uncertainty :- As per traceability	± 0.15 °C
Calibration Validity :- 28.09.2022	10.06.2022

<b>Discipline</b> :- Thermal Calibration	<b>Group</b> Temperature
--	--------------------------

<b>4.Calibration Results</b> :-				
Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ±°C
-50	-49.74	-49.4	0.34	0.40
100	100.34	100.0	0.34	0.40
300	300.56	301.3	0.74	0.40

**Note:**

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- 6) The Standard used are traceable to National / International Standard.
- 7) Conformity with requirements :- ----

Calibrated By   
Calibration Engineer  
Divya Dagadu



Approved By   
Technical Manager  
Sachin A. Mhasawade







# RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,  
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170  
Cell : 7774055755, 7774055855, 7774058855, 7774022900  
Email : reliable1010@gmail.com/reliabletechnocare@gmail.com  
Web. : www.reliable.world



## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b>	:-	Page No.	:- 1 of 1
<b>Viloo Poonawalla Memorial Hospital</b>		NABL Accreditation No.	:- CC-2927
S.No.156, Plot No.1/3A+3B/1+2, 3, Pune - Solapur Rd,		Certificate No.	:- 22.03.24.007
Near Savli Corner, Hadapsar, Pune - 411028		Date of Received	:- 24.03.2022
Ambient Temp. (°C)	:- 22.6	Date of Calibration	:- 24.03.2022
Relative Humidity (%RH)	:- 51	Next Calibration Due On	:- 23.03.2023
Location of calibration	:- In Lab	Calibration method No.	:- RTS-WI-10
Condition of Item	:- Ok	Date of Issue	:- 26.03.2022
		ULR No.	:- CC292722000002872F

<b>2. Description of Item</b>			
Name	:- Thermometer With Soncor	Range	:- -50 to 300 °C
Id No	:- VPMI I/VTI/TM/558	Least Count	:- 0.1 °C
Make	:- Multi/ST-9283 B		
Type	:- Digital		


<b>3.Detail of Equipment used for calibration</b>			
Name	:- 6½ DMM	PT 100 Sensor	
Certificate No.	:- VI/21-22/0472-03	TL/021/531.1.1	
Certified By	:- VIPL	Tempsens	
ID/Sr No	:- TW0000508A	TPRT-1574	
Range	:- As per traceability	-200 to 500 °C	
Uncertainty	:- As per traceability	± 0.15 °C	
Calibration Validity	:- 28.09.2022	10.06.2022	


**Discipline** :- Thermal Calibration **Group** Temperature


<b>4.Calibration Results</b>				
Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ±°C
-50	-49.82	-49.6	0.22	0.40
100	100.25	100.6	0.35	0.40
300	300.41	301.0	0.60	0.10

**Note:**

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- 6) The Standard used are traceable to National / International Standard.
- 7) Conformity with requirements :- ----

Calibrated By   
 Calibration Engineer  
 Divya Dagadu



Approved By   
 Technical Manager  
 Sachin A. Mhasawade







# Calibration Certificate



IC-2733

FORMAT NO. BELZ/F28

Page No. 1 of 1

Name & Address of Client		Certificate No.	30045632
M/s. Reliable Technical Services Pvt. Ltd. (A Division of Reliable Technical Services Pvt. Ltd.) Private Limited, Reliance House, 97, 7/334-35, Sarv Tularan, Mayapuri, New Delhi-110021		IR No.	CC273321300011961F
Instrument Name		Date of Issue	30/10/2021
Equipment/ Machine		Date of Calibration Start & End	27/10/2021 - 28/10/2021
Equipment Location		Recommended Date Date	28/10/2021
Make/Model		Service request No & Date	SO21-11568 - 23/10/2021
Sr. No.		Date	
61319921		28/10/2021	

Instrument Details		Instrument Name	Thermo Hygrometer With Sensor	Range/Spec	-40 to 60°C to 100%RH
Equipment/ Machine		Equipment Location	Not Specified	Accuracy/Acceptance Criteria	Not Specified
Make/Model		Sr. No.	Yudhan/Rotronic	Lead Count/Resolution	0.1°C/0.1%RH
Sr. No.		Calibrated At	61319921	Lab	
ID No.		RTS-HYG-04			

Master Equipment Used (Reference Standard/Working Standard)					
Sr. No.	Instrument Name	ID/Sr. No.	Accuracy/Acceptance Criteria	Certificate No.	Calibrated By Date
1	Super-DAQ Precision Temperature Scanner	BELZ/TDQ-AQ-01/ H358003645660220	As Per Specification	AIP/L2020/0699	Anshuman (NABL, CC-2020) 05/12/2021
2	Sold Platinum Resistance Thermometer-SPRT	BELZ/TS/SPRT-01/ 05540	+0.007°C	FL/C/TH/10032021 C001	Parag Labs (NABL, CC-2739) 19/04/2022
3	Digital RH Indicator With Sensor	BELZ/THH-03/6220(K/mh),20365987(sam)	+0.3%RH	20121146/D1.03/C-269	NPL 11/01/2022
Environmental Condition		Temperature(°C)	15±3	Calibration Method	CI(T)-07
		Relative Humidity(%RH)	35 to 70		

Calibration Results			
S.No.	Std. Avg. Readings in IN Deg C AT 50%RH	UUC Avg. Readings in IN Deg C AT 50%RH	Expanded Uncertainty(±)
01.	0.175	0.1	±0.20°C
02.	9.918	9.8	
03.	20.257	20.1	
04.	40.093	39.9	
05.	60.085	59.8	
IN %RH AT 25°C			
01.	10.69	10.3	±0.8%RH
02.	30.12	29.6	
03.	50.27	49.5	
04.	70.49	69.8	
05.	94.71	93.6	

The Expanded Uncertainty of measurement was evaluated at a confidence level of 95.45% with coverage factor, k=2 for a normal probability distribution.  
UUC-Limit Under Calibration

Remarks: ...

Notes:  
1. The Client shall refer to the particular form submitted for calibration.  
2. The Certificate shall not be used for any other purpose without the written permission of Client Executive.  
3. The Client shall refer to the particular form submitted for calibration.  
4. Calibration certificates issued for weight & length & volume & temperature & humidity & force & torque & pressure & flow & viscosity & density & refractive index & pH & conductivity & gas analysis & sound & vibration & radiation & XRF has been endorsed by NPL, New Delhi.  
5. Calibration certificates issued for weight & length & volume & temperature & humidity & force & torque & pressure & flow & viscosity & density & refractive index & pH & conductivity & gas analysis & sound & vibration & radiation & XRF has been endorsed by NPL, New Delhi. Wherever the reference is made to NPL, it shall be understood that the reference is made to NPL, New Delhi.

Calibrated by:   
Checked by:   
Approved by:   
Quality/Technical Manager



Calibration Certificate No.: TL/021/531.1.2  
 ULR No.: CC28402100000394F

Page 1 of 1

M/S RELIABLE TECHNICAL SERVICES  
 (A DIVISION OF RELIABLE TECHNOCARE PRIVATE LIMITED)  
 RELIABLE HOUSE, 207, 4334-35,  
 SANTI TUKA, 200, KANAKPUR, PUNE-411018  
 INDIA

J.O.N.O. Received Date: 10/16/2021  
 Calibration Date: 11/16/2021  
 Next Call Due On: 10/16/2022  
 Issue Date: 11/16/2021



Sample Identification: -196, -80, -35, 0, 50, 150, 250, 350, 400 & 500 °C  
 Calibration Standard: PT-100X1 SIMPLEX, 4 WIRE RTD  
 Sr. No.: 19640  
 I.D. No.: RTS-RTD-01  
 Range: -200 to 500°C  
 Location: At Lab  
 Condition of Item: Satisfactory

Calibration Procedure: The UIC was allowed to be stable at specified temperature points. The reading of Standard and UIC was recorded. This Certificate Format No. -TLAB-6. The report provides the reading and deviation. As per work instruction no. V1-LAB-36 Temperature (in °C) against RTD reading (in Ohms) are calculated by using ASTM E-1137/E1137M-04. Minimum Immersion depth for RTD is 100mm.

The UIC has been compared with the following working standards of sensor & measuring instrument.

Standards	Traceable To	Certificate No.	Valid Up to
SSPRT	TEMPSENS, UDAPUR	TL/021/285.2.1	14.03.2022
PRECISION THERMOMETER	YMPU, UDAPUR	YMPU/319215/109614	02.07.2021
CALSYS-69/50			
CALSYS 300			
CALSYS 650			

The Standards used for calibration is traceable to National / International Standard.  
 The results of calibration are as under:

Sr. No.	Standard Sensor °C	UUC		Dev. %	Expanded uncertainty (±°C)
		Ohms	°C		
1	-195.76	20.492	-195.43	0.23	0.08
2	-79.92	65.442	-79.71	0.21	0.08
3	-34.96	95.322	-34.80	0.15	0.08
4	0.00	120.095	0.14	0.03	0.08
5	49.93	119.439	50.11	0.13	0.08
6	149.68	157.377	150.14	0.25	0.08
7	249.80	194.167	250.19	0.32	0.10
8	349.75	229.807	350.26	0.5*	0.15
9	399.73	247.236	400.40	0.6*	0.15
10	499.66	281.151	500.52	0.8*	0.15

Temperature Scale: ITS -90  
 Ambient Temp.: 25±2°C  
 Relative Humidity: 50±20%RH

Calibrated By: *[Signature]*  
 Checked By: *[Signature]*  
 Technical Manager

Note:  
 (1) The calibration results reported in this certificate are valid at the time of and under the stated condition of measurement for the particular sample.  
 (2) The report shall not be retracted except in full, without written permission of the laboratory.  
 (3) Coverage factor "k" = 2 at approx 95.45% confidence level.  
 (4) Next Calibration Due Date mentioned as per customer requirement.





# VAIDYANATHESHWARA INSTRUMENTS CERTIFICATE OF CALIBRATION



COMMITTED TO THE CUSTOMER SINCE - 1986  
No. 30/1A, 4th Main Road, 3rd Cross, Rajaji Gandhi Nagar, JB Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Contact : 080-2337266, Mob : 988656789 / 9448090177 / 9894308118 | Email : info@vpigroup.com Web : www.vpigroup.com  
NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With valid Certificate No. CC-2473

1 Name and Address of Customer  
M/s. Reliable Technical services,  
(A Division Of Reliable Technocare Pvt.Ltd.)  
Reliable House 497/2834, Behind Police Station,  
Sant Tukaram Nagar, Pimpri, Pune - 411018.

2 Customer Reference  
2.1 ULR No : CC247321200000130F  
2.2 SRF No : 0472  
2.3 Form No : VI-FRM-ET-008  
2.4 Certificate No. : VI21-22/0472-02  
2.5 DC No & Calc : RTS/4521-22 & 15-09-2021  
2.6 Date of Issue : 07-10-2021

3 Details Of Device Under Calibration(DUC)  
3.1 Nomenclature : 6.5 Digit Multimeter  
3.2 Make / Model : Pincleat M3500A  
3.3 SI. No. / Id.No. : TW00006839 / RTS-DM-02  
3.4 DUC Condition : Satisfactory  
3.5 Calibration Procedure : SRF-97-08  
3.6 No. of Pages : 4  
3.7 Calibration Date : 28-09-2021  
3.8 Calibration Due : 28-09-2022  
3.9 Calibration Incent : VI Electrical Lab  
4.0 Discipline : Electro - Technical

4 Environmental Condition  
Temperature : 25.2 °C Humidity : 51 %RH

5 Standards Used for Calibration

SI. No.	Nomenclature	Make & Model	SI No	Basic Accuracy	Certificate No.	Validity
1	Calibrator	Fluke & 5522A	3289904	±11 ppm	CRP/CAL/51653/1	14-09-2022

- 6 Note:
- The Calibration Certificate relates only to the above DUC.
  - Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
  - Corrections/changes, invalidate the Calibration Certificate.
  - Calibration of the DUC are traceable to National standards/International Standards
  - Any error in this Certificate should be brought to our knowledge within 30 days from the date of the Certificate.
  - Results Reported are valid at the time of and under the stated conditions of measurements
  - The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Checked By  
Navya Hemant  
(Lab In-Charge)



This report refers only to the sample submitted and may not be reproduced except in full without written permission from Vaidyanatheshwara Instruments Bangalore - 560096



# VAIDYANATHESHWARA INSTRUMENTS CERTIFICATE OF CALIBRATION



COMMITTED TO THE CUSTOMER SINCE - 1986  
No. 30/1A, 4th Main Road, 3rd Cross, Rajaji Gandhi Nagar, JB Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Contact : 080-2337266, Mob : 988656789 / 9448090177 / 9894308118 | Email : info@vpigroup.com Web : www.vpigroup.com  
NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With valid Certificate No. CC-2473

Certificate No. : VI21-22/0472-02

Results:

Sl. No.	Parameter	Range	Standard T-put	DUC Reading	Error Claimed (%)	Deviation Observed	Measurement Uncertainty (%)	k
1	DC Current	100 mA	1D	0.5003 mV	0.0035 mV	0.0003 mV	0.0012 mV	2.0
2			2E	10.0005	0.0040	0.0005	0.00039	2.0
3			3E	20.0010	0.0045	0.0010	0.00078	2.0
4			3E	25.0012	0.0048	0.0012	0.00098	2.0
5			3E	30.0013	0.0050	0.0013	0.00117	2.0
6			3E	35.0018	0.0053	0.0018	0.00085	2.0
7			4E	40.0020	0.0055	0.0020	0.00158	2.0
8			5E	50.0023	0.0060	0.0023	0.00195	2.0
9			120	100.0025	0.0065	0.0025	0.00350	2.0
10			10	0.100005 V	0.000011 V	0.000003 V	0.000004 V	2.0
11	DC Voltage	1 V	C3	0.500012	0.000027	0.000012	0.000008	2.0
12			1	1.500022	0.000047	0.000022	0.000015	2.0
13			1V	1.00003 V	0.00009 V	0.00003 V	0.000015 V	2.0
14	10 V	10 V	3	5.00011	0.00023	0.00011	0.000090	2.0
15			TC	10.00019	0.00040	0.00019	0.000160	2.0
16			TC	10.0004 V	0.0011 V	0.0004 V	0.000150	2.0
17	100 V	100 V	5E	50.0010	0.00029	0.00010	0.0001150	2.0
18			TC	100.0013	0.0051	0.0013	0.002300	2.0
19			100V	100.003 V	0.015 V	0.003 V	0.002300 V	2.0
20	1000 V	1000 V	50E	500.011	0.033	0.011	0.011500	2.0
21			1120	999.984	0.055	-0.016	0.023000	2.0
22			0	0.10017 mA	0.00017 mA	0.00017 mA	0.00002 mA	2.0
23	10 mA	10 mA	E	5.00041	0.00450	0.00041	0.00035	2.0
24			13	10.00102	0.03700	0.00102	0.00110	2.0
25			10 mA	10.0019 mA	0.0100 mA	0.0019 mA	0.00110 mA	2.0
26	100 mA	100 mA	50	50.0038	0.0300	0.0038	0.00550	2.0
27			TC	100.0049	0.0550	0.0049	0.01100	2.0
28			0.1A	0.100015 A	0.000200 A	0.000015 A	0.00001 A	2.0
29	1 A	1 A	C3	0.800102	0.003500	0.000102	0.00014	2.0
30			1	1.000203	0.001100	0.000203	0.00027	2.0
31			1A	1.000025 A	0.00180 A	0.000025 A	0.00027 A	2.0
32	3 A	3 A	2	2.00032	0.00300	0.00032	0.00035	2.0
33			3	3.00123	0.00420	0.00123	0.00035	2.0

Checked By  
Navya Hemant  
(Calibration Engineer)

Checked By  
Navya Hemant  
(Lab In-Charge)



This report refers only to the sample submitted and may not be reproduced except in full without written permission from Vaidyanatheshwara Instruments Bangalore - 560096